

Brunswick County Public Utilities - NC

PO Box 249
Bolivia, NC 28422-0249

LELAND N.C.

Client Project# NORTHWEST WATER PLANT
Samples Received: 2/14/2025

Analytical Report 0225-819

PFAS by Isotope Dilution (non-potable water)

Report Issue Date: 3/12/2025

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This analytical report was prepared in Portable Document Format (.PDF) and contains 27 pages. This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:



Laura Boivin, QA Associate II



Enthalpy Analytical, LLC – Wilmington
Christina Kurnath, Project Manager
chkurnath@montrose-env.com / www.enthalpy.com
O: 910-876-6895
2714 Exchange Drive, Wilmington, NC 28405

Table of Contents

Case Narrative	
General Reporting Notes	
PFAS Acronym List.....	
Results	
Summary of Results	
Detailed Results	
QC Data	
Method Blank.....	
LCS.....	
Sample Custody.....	
Chain of Custody	

Narrative Summary

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0225-819-1
Client ID.	NORTHWEST WATER PLANT Site: LELAND N.C.

1. Custody

Meredith Curtis received the samples at 1.3 °C after being relinquished by Brunswick County Public Utilities - NC.

The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC.

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix	Received
0225-819-001-1	021425-SO1	aqueous	2025-02-14
0225-819-001-2	021425-SO1	aqueous	2025-02-14
0225-819-002-1	021425-EO1	aqueous	2025-02-14
0225-819-002-2	021425-EO1	aqueous	2025-02-14

2. Methods and Analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU047	Brunswick List	ENVI-Carb

3. Analysis

The samples were analyzed using Sciex Triple Quad 7500 (LC/MS/MS "Bumblebee") and using Waters Acquity UPLC equipped with Xevo TQ MS (LC/MS/MS "Pippin").

The samples were analyzed using more than one batch preparation to include all analytes of interest.

4. Calibration

In the initial calibration, the reported analytes exhibited R^2 of = 0.99. The reported analytes in the calibration standards, Initial Calibration Verification (ICV) and continuing calibration (concal) met the accuracy criterion for native analytes, except as noted below.

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0225-819-1
Client ID.	NORTHWEST WATER PLANT Site: LELAND N.C.

The Standards that did not were:

- SID BH58 (10:2 FTS)
- SID BH67 (10:2 FTS)

Analyte(s) that exceeded method control limits in the concals were not detected >LOQ in the samples. The data is reported without adverse impact.

5. QC Notes

Ongoing Precision Recovery (OPR) control limits have not been established for some analytes of interest.

Except where noted below, the QC sample analyses passed all method criteria.

QC samples that did not meet method acceptance criteria were:

- MB_19032_PFAAS (M2-8:2 FTS, M2PFTeDA, d3-N-MeFOSA, d5-N-EtFOSA)
- OPR_19032_PFAAS (PFTTrDA, M2-8:2 FTS, M2PFTeDA, d3-N-MeFOSA, d5-N-EtFOSA)

PFTTrDA was detected above method control limits in the Ongoing Precision Recovery (OPR) QC sample. Impacted analyte was not detected >LOQ in the samples; therefore, the data is reportable with no adverse impact.

See Reporting Notes below.

PFAS by Isotope Dilution (non-potable water) samples were extracted within 28 days, and extracts analyzed within 28 days.

6. Reporting Notes

The results presented in this report are representative of the samples as provided to the laboratory.

This report provides all results including detections below LOD following client instruction.

Some labeled extraction standards (ES) in the analyses recovered outside method control limits, as denoted by the "Q" qualifier. The target analytes are quantified based on their ratio to their labeled standard analogs. As a result, low or high labeled standard recovery do not cause any change to ratios or contribute any additional error in the measurement of the target analytes. When detected at a signal-to-noise above 10:1 the ES peak area is used to quantify its respective target analyte using accepted isotope dilution principles. The data is reported without adverse impact.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2016 TNI Standard under certificate number 05075.



General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, LLC - Wilmington, NC data reports, unless specifically noted otherwise.

General Data Qualifiers

- Ac - Alternate calculation flag indicates the es recovery was calculated using the opening concal when either of the following situations is encountered in the data processing software: the ES recovery is over 400% or the JS is not detected.
- B – The analyte was found in the method blank, at a concentration that was at least 10% of the amount in the sample.
- Cxx – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group ('xx') are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve). For HRMS data, this condition does not imply additional measurement uncertainty. For LC-MS/MS data, these values should be considered as having measurement uncertainty higher than values within the calibration range.
- EDL – Estimated Detection Level: The EDL is unique to isotope dilution methods and reflects the conditions of analysis at the time of analysis, including the equipment used. Where the MDL is a static value, the EDL is a dynamic value.
- EMPC – Estimated Maximum Possible Concentration: EMPC is specific to Dioxin/Furan tests to indicate the determined ion-abundance ratio was outside the allowed theoretical range (usually due to being near the detection limit, although it can very rarely be caused by a co-eluting interference). The EMPC concentration is adjusted to reflect the value at the theoretical ion-abundance ratio.
- I/IR – The ion ratio between the primary and secondary ions was observed to be outside the method criteria. The analyte concentration may be inaccurate due to interference.
- J – The analyte has a concentration below the minimum calibration level (LOQ value) but greater than the LOD. These values should be considered as having measurement uncertainty higher than values within the calibration range
- L - For reports containing PFAS analytes only, this flag indicates that an analyte has a concentration below the Minimum Detection Limit (MDL) . The reported concentration is not recommended for regulatory use as the analyte signal may have a signal-to-noise ratio less than the criteria deemed necessary to be considered a detected analyte.
- LOD – Limit of Detection: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOD. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the minimum detection limit (MDL). The LOD is adjusted for sample weight or volume.



General Reporting Notes – Data Qualifiers

- LOQ – Limit of Quantitation: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOQ. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the reporting limit (RL). The LOQ is adjusted for sample weight or volume.
- <LOD() – Analyte was not found at a concentration high enough to be reported as detected. It is reported as less than the LOD, and the LOD is given in the parentheses.
- <LOQ() – Analyte was not found at a concentration high enough to be reported as above the QSM-defined LOQ or TNI defined Reporting Limit. It is reported as less than the LOQ, and the LOQ is given in the parentheses.
- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable due to issues observed in sample preparation or analysis.
- PR – The associated congener(s) is(are) poorly resolved.
- QI – Indicates the presence of a quantitative interference.
- RL – Reporting Limit. Lowest reportable value. The level is higher than the MDL.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected.
- V / Q – The labeled standard recovery is not within method control limits.
- X – Indicates the result is from re-injection/repeat/second-column analysis.

Lab Identifiers/ Data Attributes

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Dilution Data. Result was obtained from the analysis of a dilution. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.



General Reporting Notes – Data Qualifiers

- R – Indicates a re-extraction of the sample.
- RJ – Indicates a reinjection of the sample extract.
- S – Indicates a sample split. The number that follows the “S” indicates the split factor.
- SAT – Indicates an analyte saturated the detector.

PFAS Compound Acronym List			
Acronym	CAS #	Compound Name	
* accredited for SOP EU047 / EPA method 1633 # Method 537.1 Accredited ^ Method 533 Accredited ~EPA 1633 extended list			
Target Analytes			
* , ^	PFBA	375-22-4	Perfluorobutanoic Acid
* , # , ^	PFPeA	2706-90-3	Perfluoropentanoic Acid
* , # , ^	PFHxA	307-24-4	Perfluorohexanoic Acid
* , # , ^	PFHpA	375-85-9	Perfluoroheptanoic Acid
* , # , ^	PFOA	335-67-1	Perfluorooctanoic Acid
* , # , ^	PFNA	375-95-1	Perfluorononanoic Acid
* , # , ^	PFDA	335-76-2	Perfluorodecanoic acid
* , # , ^	PFUnA (PFUnDA)	2058-94-8	Perfluoroundecanoic acid
* , #	PFDoA (PFDoDA)	307-55-1	Perfluorododecanoic acid
* , #	PFTriA (PFTriA)	72629-94-8	Perfluorotridecanoic acid
* , # , ^	PFTeDA (PFTA)	376-06-7	Perfluorotetradecanoic acid
* , ^	PFBS	375-73-5	Perfluorobutane sulfonic acid
* , # , ^	PFPeS	2706-91-4	Perfluoropentane sulfonic acid
* , ^	PFHxS	355-46-4	Perfluorohexane sulfonic acid
* , # , ^	PFHpS	375-92-8	Perfluoroheptane sulfonic acid
* , # , ^	PFOS	1763-23-1	Perfluorooctane sulfonic acid
* , #	PFNS	68259-12-1	Perfluorononane sulfonic acid
* , #	PFDS	335-77-3	Perfluorodecane sulfonic acid
* , ^	4:2 FTS	757124-72-4	4:2 fluorotelomer sulfonic acid
* , ^	6:2 FTS	27619-97-2	6:2 fluorotelomer sulfonic acid
* , ^	8:2 FTS	39108-34-4	8:2 fluorotelomer sulfonic acid
~	10:2 FTS	120226-60-0	Fluorotelomer sulfonate 10:2
~	FHxSA	41997-13-1	Perfluorohexanesulfonamide
* , #	PFOSA (FOSA)	754-91-6	Perfluorooctane sulfonamide
* , #	N-MeFOSAA	2355-31-9	N-methyl perfluorooctane sulfonamido acetic acid
* , #	N-MeFOSA	31506-32-8	N-methylperfluoro-1-octanesulfonamide
* , #	N-MeFOSE	24448-09-7	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol
* , #	N-EtFOSAA	2991-50-6	N-ethyl perfluorooctane sulfonamido acetic acid
* , #	N-EtFOSA	4151-50-2	N-ethylperfluoro-1-octanesulfonamide
* , #	N-EtFOSE	1691-99-2	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol
* , # , ^	HFPO-DA	13252-13-6	2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy)-propanoic acid (Gen-X)
* , # , ^	11Cl-PF3OUdS	763051-92-9	11-chloroheptafluoro-3-oxadecane-1-sulfonic acid
* , # , ^	9Cl-PF3ONS	756426-58-1	9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid
* , # , ^	ADONA	919005-14-4	4,8-dioxa-3H-perfluorononanoic acid
* , ^	PFEESA	113507-82-7	Perfluoro(2-ethoxyethane)sulphonic acid
* , ^	PFMOBA (PFMBA)	863090-89-5	Perfluoro-4-methoxybutanoic acid
* , ^	NFDHA	151772-58-6	Nonafluoro-3,6-dioxahexanoic acid
* , ^	PFMOPrA (PFMPA)	377-73-1	Perfluoro-3-methoxypropanoic acid
~	PFPrA	422-64-0	2,2,3,3,3-Pentafluoropropionic acid
~	PFPrS (PFPS)	423-41-6	Perfluoropropanesulfonic acid
~	PFMOAA	674-13-5	Perfluoro-2-methoxyacetic acid
~	PFO2HxA	39492-88-1	Perfluoro (3,5-dioxahexanoic) acid
~	PFO3OA	39492-89-2	Perfluoro (3,5,7-trioxaoctanoic) acid
~	PFO4DA	39492-90-5	Perfluoro (3,5,7,9-tetraoxadecanoic) acid
~	PFO5DA	39492-91-6	Perfluoro(3,5,7,9,11-pentaoxadodecanoic) acid
~	Nafion Byproduct 1 (PS Acid)	29311-67-9	Nafion Byproduct 1
~	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	Nafion Byproduct 2
~	PEPA	267239-61-2	Perfluoro-2-ethoxypropanoic acid
~	PMPA	13140-29-9	Perfluoro-2-methoxypropanoic acid

PFAS Compound Acronym List		
Acronym	CAS #	Compound Name
* accredited for SOP EU047 / EPA method 1633 # Method 537.1 Accredited ^ Method 533 Accredited ~EPA 1633 extended list		
~ PFECA-G	801212-59-9	4-(Heptafluoroisopropoxy)hexafluorobutanoic acid
~ PFHxDA	67905-19-5	Perfluorohexadecanoic acid
~ R-PSDA (Nafion Byproduct 4)	2416366-18-0	Perfluoro-4-(2-sulfoethoxy)pentanoic acid
Hydrolyzed PSDA (Nafion Byproduct 5)	2416366-19-1	2-fluoro-2-[1,1,2,3,3,3-hexafluoro-2-(1,1,2-tetrafluoro-2-sulfoethoxy)propoxy]-acetic acid
~ R-PSDCA (Nafion Byproduct 6)	2416366-21-5	1,1,2,2-tetrafluoro-2-[1,2,2,3,3-pentafluoro-1-(trifluoromethyl)propoxy] ethanesulfonic acid
~ EVE Acid	69087-46-3	2,2,3,3-tetrafluoro-3-({1,1,1,2,3,3-hexafluoro-3-[(1,2,2-trifluoroethenyl)oxy]propan-2-yl)oxy}propionic acid
~ FBSA	30334-69-1	Perfluorobutylsulfonamide
~ MeFBSA	68298-12-4	1-Butanesulfonamide; (N-(Methyl)nonafluorobutanesulfonamide)
~ Hydro-EVE Acid	773804-62-9	2,2,3,3-Tetrafluoro-3-[[1,1,1,2,3,3-hexafluoro-3-(1,2,2,2-tetrafluoroethoxy)propan-2-yl]oxy}propanoic acid
~ R-EVE Acid	2416366-22-6	4-(2-carboxy-1,1,2,2-tetrafluoroethoxy)-2,2,3,3,4,5,5,5-octafluoro-pentanoic acid
~ NVHOS	1132933-86-8	Perfluoroethoxysulfonic acid
*~ PFDoS	79780-39-5	Perfluorododecane sulfonic acid
~ PFODA	16517-11-6	Perfluorooctadecanoic acid
* 3:3 FTCA	356-02-5	2H,2H,3H,3H-Perfluorohexanoic acid
* 5:3 FTCA	914637-49-3	2H,2H,3H,3H-Perfluorooctanoic acid
* 7:3 FTCA	812-70-4	2H,2H,3H,3H-Perfluorodecanoic acid
~ N-AP-FHxSA	50598-28-2	N-(3-(Dimethylamino)propyl)tridecafluoro-1-hexanesulfonamide
~ N-CMAmP-6:2 FOSA	34455-29-3	N-(Carboxymethyl)-N,N-dimethyl-3-(((3,3,4,4,5,5,6,6,7,7,8,8,8-tridecafluorooctyl)sulfonyl)amino)1-propanaminium
~ BPAF	1478-61-1	Bisphenol AF
~ HQ-115	90076-65-6	Bis(trifluoromethane)sulfonimide lithium salt

Results

Enthalpy Analytical

Job No.: 0225-819-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Summary

	Compound	CAS	021425-SO1 ng/L	021425-EO1 ng/L
Acids	PFPrA	422-64-0	514 L	440 L
	PFBA	375-22-4	3.51	3.70
	PFPeA	2706-90-3	7.02	7.37
	PFHxA	307-24-4	7.24	7.76
	PFHpA	375-85-9	3.78	3.79
	PFOA	335-67-1	7.15	7.85
	PFNA	375-95-1	0.714	0.753
	PFDA	335-76-2	0.205 J	0.184 J
	PFUnDA	2058-94-8	ND U	ND U
	PFDoDA	307-55-1	ND U	ND U
	PFTrDA	72629-94-8	ND U	ND U
	PFTeDA	376-06-7	ND U	ND U
	PFHxDA	67905-19-5	ND U	ND U
	Sulfonates	PFBS	375-73-5	3.47
PFPeS		2706-91-4	0.824	0.686
PFHxS		355-46-4	4.14	4.13
PFHpS		375-92-8	0.213 L	0.226 L
PFOS		1763-23-1	10.7	10.8
PFNS		68259-12-1	ND U	ND U
PFDS		335-77-3	ND U	ND U
4:2 FTS		757124-72-4	ND U	ND U
6:2 FTS		27619-97-2	0.0720 L	0.111 L
8:2 FTS		39108-34-4	ND U	ND U
10:2 FTS	120226-60-0	ND U	ND U	
Sulfonamidos	FBSA	30334-69-1	0.302 J	0.263 L
	N-EtFOSA	4151-50-2	ND U	ND U
	N-EtFOSAA	2991-50-6	ND U	ND U
	N-EtFOSE	1691-99-2	ND U	ND U
	N-MeFOSA	31506-32-8	ND U	ND U
	N-MeFOSAA	2355-31-9	ND U	ND U
	N-MeFOSE	24448-09-7	ND U	ND U
	PFOSA	754-91-6	2.18	0.0668 L
	PFECAs	ADONA	919005-14-4	ND U
EVE Acid		69087-46-3	ND U	ND U
HFPO-DA		13252-13-6	2.66	2.33
Hydro-EVE Acid		773804-62-9	0.0381 L	0.0442 L
NFDHA		151772-58-6	ND U	ND U
PEPA		267239-61-2	1.48	1.15
PFECA-G		801212-59-9	ND U	ND U
PFMOAA		674-13-5	14.1	13.5
PFMOBA		863090-89-5	ND U	ND U
PFMOPrA		377-73-1	0.0447 L	0.0359 L
PFO2HxA		39492-88-1	2.66	2.35
PFO3OA		39492-89-2	0.751	0.657
PFO4DA		39492-90-5	0.129 L	0.112 L
PFO5DA		39492-91-6	ND U	ND U
PMPA		13140-29-9	4.25	3.26
R-EVE		2416366-22-6	9.33	9.14
PFESAs	11Cl-PF3OUds	763051-92-9	ND U	ND U
	9Cl-PF3ONS	756426-58-1	ND U	ND U
	Hydrolyzed PSDA	2416366-19-1	2.75	2.28
	Nafion Byproduct 1 (PS Acid)	29311-67-9	ND U	ND U
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	0.149 L	0.0852 L
	NVHOS	1132933-86-8	ND U	ND U
	PFEESA	113507-82-7	ND U	ND U
	R-PSDA	2416366-18-0	5.27	4.86
	R-PSDCA	2416366-21-5	ND U	ND U

Enthalpy Analytical

Job No.: 0225-819-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	021425-SO1		
Sampling Site			
Enthalpy ID	0225-819-001-1	Prep Batch	eu18976
Matrix	aqueous	Analyst	ext-magennaef
Sampling Date	2025-02-14 13:05	Instrument	Bumblebee
Received Date	2025-02-14	Sample Vol mL	0.1
Prep Date	2025-02-18 12:15	Extract Vol mL	0.2
AnalysisDate	2025-02-21 14:53	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	A		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	B210225-02211453	514	700	1530			L
ES	13C3-PFPrA		B210225-02211453				20-150%	100%	

Enthalpy Analytical

Job No.: 0225-819-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	021425-SO1	Prep Batch	EU19032
Sampling Site		Analyst	jogres
Enthalpy ID	0225-819-001-2	Instrument	Pippin
Matrix	aqueous	Sample Vol mL	282.28
Sampling Date	2025-02-14 13:05	Extract Vol mL	0.4
Received Date	2025-02-14	Split Factor	N/A
Prep Date	2025-02-28 09:17	Method Code	EU-047-NPW
AnalysisDate	2025-03-03 20:46		
SampleType	Sample		
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	P030325009	3.51	0.225	0.567			
	PFPeA	2706-90-3	P030325009	7.02	0.162	0.567			
	PFFhxA	307-24-4	P030325009	7.24	0.190	0.567			
	PFFHpA	375-85-9	P030325009	3.78	0.198	0.567			
	PFOA	335-67-1	P030325009	7.15	0.130	0.567			
	PFNA	375-95-1	P030325009	0.714	0.128	0.567			
	PFDA	335-76-2	P030325009	0.205	0.162	0.567			J
	PFUnDA	2058-94-8	P030325009	ND	0.128	0.567			U
	PFDoDA	307-55-1	P030325009	ND	0.230	0.567			U
	PFTrDA	72629-94-8	P030325009	ND	0.188	0.567			U
	PFTeDA	376-06-7	P030325009	ND	0.216	0.567			U
	PFFhxDA	67905-19-5	P030325009	ND	0.301	0.567			U
	Sulfonates	PFBS	375-73-5	P030325009	3.47	0.301	0.567		
PFPeS		2706-91-4	P030325009	0.824	0.116	0.534			
PFFhXS		355-46-4	P030325009	4.14	0.438	0.519			
PFFHpS		375-92-8	P030325009	0.213	0.275	0.540			L
PFOS		1763-23-1	P030325009	10.7	0.299	0.525			
PFNS		68259-12-1	P030325009	ND	0.176	0.546			U
PFDS		335-77-3	P030325009	ND	0.298	0.546			U
4:2 FTS		757124-72-4	P030325009	ND	0.0735	0.531			U
6:2 FTS		27619-97-2	P030325009	0.0720	0.267	0.540			L
8:2 FTS		39108-34-4	P030325009	ND	0.127	0.543			U
10:2 FTS	120226-60-0	P030325009	ND	0.434	0.567			U	
Sulfonamidos	FBSA	30334-69-1	P030325009	0.302	0.269	0.567			J
	N-EiFOSA	4151-50-2	P030325009	ND	0.351	0.567			U
	N-EiFOSAA	2991-50-6	P030325009	ND	0.230	0.567			U
	N-EiFOSE	1691-99-2	P030325009	ND	0.868	2.55			U
	N-MeFOSA	31506-32-8	P030325009	ND	0.234	0.567			U
	N-MeFOSAA	2355-31-9	P030325009	ND	0.159	0.567			U
	N-MeFOSE	24448-09-7	P030325009	ND	0.538	2.55			U
	PFOSA	754-91-6	P030325009	2.18	0.0795	0.567			
	ADONA	919005-14-4	P030325009	ND	0.154	0.537			U
	EVE Acid	69087-46-3	P030325009	ND	0.181	1.28			U
PFECAs	HFPO-DA	13252-13-6	P030325009	2.66	0.0600	0.567			
	Hydro-EVE Acid	773804-62-9	P030325009	0.0381	0.186	0.567			L
	NFDHA	151772-58-6	P030325009	ND	0.119	0.567			U
	PEPA	267239-61-2	P030325009	1.48	0.106	0.567			
	PFECA-G	801212-59-9	P030325009	ND	0.0756	0.567			U
	PFMOAA	674-13-5	P030325009	14.1	0.287	0.567			
	PFMOBA	863090-89-5	P030325009	ND	0.951	1.28			U
	PFMOPrA	377-73-1	P030325009	0.0447	0.202	0.567			L
	PFO2HxA	39492-88-1	P030325009	2.66	0.182	0.567			
	PFO3OA	39492-89-2	P030325009	0.751	0.260	0.567			
	PFO4DA	39492-90-5	P030325009	0.129	0.448	2.83			L
	PFO5DA	39492-91-6	P030325009	ND	0.453	2.83			U
	PMPA	13140-29-9	P030325009	4.25	0.134	0.567			
	R-EVE	2416366-22-6	P030325009	9.33	0.941	1.28			
	PFESAs	11CI-PF3OUdS	763051-92-9	P030325009	ND	0.267	0.534		
9CI-PF3ONS		756426-58-1	P030325009	ND	0.363	0.528			U
Hydrolyzed PSDA		2416366-19-1	P030325009	2.75	0.377	0.567			
Nafion Byproduct 1 (PS Acid)		29311-67-9	P030325009	ND	0.303	0.567			U
Nafion Byproduct 2 (Hydro-PS Acid)		749836-20-2	P030325009	0.149	0.469	0.567			L
NVHOS		1132933-86-8	P030325009	ND	0.0873	0.567			U
PFEESA		113507-82-7	P030325009	ND	0.170	0.567			U
R-PSDA		2416366-18-0	P030325009	5.27	2.50	2.50			
R-PSDCA		2416366-21-5	P030325009	ND	0.239	0.567			U
ES		MPFBA		P030325009				20-150%	93.8%
	M5PFPeA		P030325009				20-150%	215%	Q
	M3PFBS		P030325009				20-150%	234%	Q
	M2-4:2 FTS		P030325009				20-150%	155%	Q
	M5PFFhxA		P030325009				20-150%	91.5%	
	M3HFPO-DA		P030325009				20-150%	74.3%	
	M4PFFHpA		P030325009				20-150%	98.5%	

Enthalpy Analytical

Job No.: 0225-819-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	021425-SO1		
Sampling Site			
Enthalpy ID	0225-819-001-2	Prep Batch	EU19032
Matrix	aqueous	Analyst	jogres
Sampling Date	2025-02-14 13:05	Instrument	Pippin
Received Date	2025-02-14	Sample Vol mL	282.28
Prep Date	2025-02-28 09:17	Extract Vol mL	0.4
AnalysisDate	2025-03-03 20:46	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	-		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M3PFHxS		P030325009				20-150%	101%	
M2-6:2 FTS		P030325009				20-150%	114%	
M8PFOA		P030325009				20-150%	89.4%	
M9PFNA		P030325009				20-150%	87.5%	
M8PFOS		P030325009				20-150%	82.5%	
M2-8:2 FTS		P030325009				20-150%	163%	Q
M8FOSA-I		P030325009				20-150%	48.6%	
M6PFDA		P030325009				20-150%	75.7%	
d3-N-MeFOSAA		P030325009				20-150%	87.9%	
d5-N-EtFOSAA		P030325009				20-150%	76.4%	
M7PFUdA		P030325009				20-150%	55.3%	
MPFDoA		P030325009				20-150%	31.1%	
M2PFTeDA		P030325009				20-150%	5.02%	Q
d3-N-MeFOSA		P030325009				10-200%	0.159%	Q
d5-N-EtFOSA		P030325009				10-200%	0.146%	Q
d7-N-MeFOSE		P030325009				10-200%	7.19%	Q
d9-N-EtFOSE		P030325009				10-200%	7.63%	Q

Enthalpy Analytical

Job No.: 0225-819-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	021425-EO1		
Sampling Site			
Enthalpy ID	0225-819-002-1	Prep Batch	eu18976
Matrix	aqueous	Analyst	ext-magennaef
Sampling Date	2025-02-14 13:05	Instrument	Bumblebee
Received Date	2025-02-14	Sample Vol mL	0.1
Prep Date	2025-02-18 12:15	Extract Vol mL	0.2
AnalysisDate	2025-02-21 15:04	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	A		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPfA	422-64-0	B210225-02211504	440	700	1530			L
ES	13C3-PFPfA		B210225-02211504				20-150%	101%	

Enthalpy Analytical

Job No.: 0225-819-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	021425-EO1	Prep Batch	EU19032
Sampling Site		Analyst	jogres
Enthalpy ID	0225-819-002-2	Instrument	Pippin
Matrix	aqueous	Sample Vol mL	283.91
Sampling Date	2025-02-14 13:05	Extract Vol mL	0.4
Received Date	2025-02-14	Split Factor	N/A
Prep Date	2025-02-28 09:17	Method Code	EU-047-NPW
AnalysisDate	2025-03-03 21:08		
SampleType	Sample		
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	P030325010	3.70	0.224	0.564				
	PFPeA	2706-90-3	P030325010	7.37	0.161	0.564				
	PFFhxA	307-24-4	P030325010	7.76	0.188	0.564				
	PFFHpA	375-85-9	P030325010	3.79	0.197	0.564				
	PFOA	335-67-1	P030325010	7.85	0.129	0.564				
	PFNA	375-95-1	P030325010	0.753	0.127	0.564				
	PFDA	335-76-2	P030325010	0.184	0.161	0.564			J	
	PFUnDA	2058-94-8	P030325010	ND	0.127	0.564			U	
	PFDoDA	307-55-1	P030325010	ND	0.229	0.564			U	
	PFTrDA	72629-94-8	P030325010	ND	0.187	0.564			U	
	PFTeDA	376-06-7	P030325010	ND	0.215	0.564			U	
	PFFhxDA	67905-19-5	P030325010	ND	0.299	0.564			U	
	Sulfonates	PFBS	375-73-5	P030325010	3.62	0.299	0.564			
		PFPeS	2706-91-4	P030325010	0.686	0.116	0.531			
PFFhXS		355-46-4	P030325010	4.13	0.435	0.516				
PFFHpS		375-92-8	P030325010	0.226	0.273	0.537			L	
PFOS		1763-23-1	P030325010	10.8	0.298	0.522				
PFNS		68259-12-1	P030325010	ND	0.175	0.543			U	
PFDS		335-77-3	P030325010	ND	0.296	0.543			U	
4:2 FTS		757124-72-4	P030325010	ND	0.0731	0.528			U	
6:2 FTS		27619-97-2	P030325010	0.111	0.266	0.537			L	
8:2 FTS		39108-34-4	P030325010	ND	0.126	0.540			U	
10:2 FTS	120226-60-0	P030325010	ND	0.431	0.564			U		
Sulfonamidos	FBSA	30334-69-1	P030325010	0.263	0.268	0.564			L	
	N-EiFOSA	4151-50-2	P030325010	ND	0.349	0.564			U	
	N-EiFOSAA	2991-50-6	P030325010	ND	0.229	0.564			U	
	N-EiFOSE	1691-99-2	P030325010	ND	0.863	2.54			U	
	N-MeFOSA	31506-32-8	P030325010	ND	0.232	0.564			U	
	N-MeFOSAA	2355-31-9	P030325010	ND	0.158	0.564			U	
	N-MeFOSE	24448-09-7	P030325010	ND	0.535	2.54			U	
	PFOSA	754-91-6	P030325010	0.0668	0.0791	0.564			L	
	ADONA	919005-14-4	P030325010	ND	0.153	0.534			U	
	EVE Acid	69087-46-3	P030325010	ND	0.180	1.27			U	
PFECAs	HFPO-DA	13252-13-6	P030325010	2.33	0.0597	0.564				
	Hydro-EVE Acid	773804-62-9	P030325010	0.0442	0.185	0.564			L	
	NFDHA	151772-58-6	P030325010	ND	0.119	0.564			U	
	PEPA	267239-61-2	P030325010	1.15	0.106	0.564				
	PFECA-G	801212-59-9	P030325010	ND	0.0752	0.564			U	
	PFMOAA	674-13-5	P030325010	13.5	0.285	0.564				
	PFMOBA	863090-89-5	P030325010	ND	0.946	1.27			U	
	PFMOPrA	377-73-1	P030325010	0.0359	0.201	0.564			L	
	PFO2HxA	39492-88-1	P030325010	2.35	0.181	0.564				
	PFO3OA	39492-89-2	P030325010	0.657	0.259	0.564				
	PFO4DA	39492-90-5	P030325010	0.112	0.446	2.82			L	
	PFO5DA	39492-91-6	P030325010	ND	0.451	2.82			U	
	PMPA	13140-29-9	P030325010	3.26	0.133	0.564				
	R-EVE	2416366-22-6	P030325010	9.14	0.935	1.27				
	PFESAs	11CI-PF3OUdS	763051-92-9	P030325010	ND	0.266	0.531			U
		9CI-PF3ONS	756426-58-1	P030325010	ND	0.361	0.525			U
Hydrolyzed PSDA		2416366-19-1	P030325010	2.28	0.375	0.564				
Nafion Byproduct 1 (PS Acid)		29311-67-9	P030325010	ND	0.301	0.564			U	
Nafion Byproduct 2 (Hydro-PS Acid)		749836-20-2	P030325010	0.0852	0.467	0.564			L	
NVHOS		1132933-86-8	P030325010	ND	0.0868	0.564			U	
PFEESA		113507-82-7	P030325010	ND	0.169	0.564			U	
R-PSDA		2416366-18-0	P030325010	4.86	2.48	2.48				
R-PSDCA		2416366-21-5	P030325010	ND	0.238	0.564			U	
ES	MPFBA		P030325010				20-150%	96.7%		
	M5PFPeA		P030325010				20-150%	188%	Q	
	M3PFBS		P030325010				20-150%	224%	Q	
	M2-4:2 FTS		P030325010				20-150%	162%	Q	
	M5PFFhxA		P030325010				20-150%	84.0%		
	M3HFPO-DA		P030325010				20-150%	67.6%		
	M4PFFHpA		P030325010				20-150%	103%		

Enthalpy Analytical

Job No.: 0225-819-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	021425-EO1		
Sampling Site			
Enthalpy ID	0225-819-002-2	Prep Batch	EU19032
Matrix	aqueous	Analyst	jogres
Sampling Date	2025-02-14 13:05	Instrument	Pippin
Received Date	2025-02-14	Sample Vol mL	283.91
Prep Date	2025-02-28 09:17	Extract Vol mL	0.4
AnalysisDate	2025-03-03 21:08	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	-		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M3PFHxS		P030325010				20-150%	101%	
M2-6:2 FTS		P030325010				20-150%	151%	Q
M8PFOA		P030325010				20-150%	95.9%	
M9PFNA		P030325010				20-150%	108%	
M8PFOS		P030325010				20-150%	97.9%	
M2-8:2 FTS		P030325010				20-150%	242%	Q
M8FOSA-I		P030325010				20-150%	79.7%	
M6PFDA		P030325010				20-150%	96.3%	
d3-N-MeFOSAA		P030325010				20-150%	122%	
d5-N-EtFOSAA		P030325010				20-150%	128%	
M7PFUdA		P030325010				20-150%	85.6%	
MPFDoA		P030325010				20-150%	77.4%	
M2PFTeDA		P030325010				20-150%	19.7%	Q
d3-N-MeFOSA		P030325010				10-200%	2.10%	Q
d5-N-EtFOSA		P030325010				10-200%	1.91%	Q
d7-N-MeFOSE		P030325010				10-200%	44.3%	
d9-N-EtFOSE		P030325010				10-200%	37.3%	

QC Data

Enthalpy Analytical

Job No.: 0225-819-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	MB_18976_PFA		
Sampling Site			
Enthalpy ID	MB_18976_PFA	Prep Batch	eu18976
Matrix	aqueous	Analyst	ext-magennaef
Sampling Date		Instrument	Bumblebee
Received Date		Sample Vol mL	0.1
Prep Date	2025-02-18 12:15	Extract Vol mL	0.2
AnalysisDate	2025-02-21 14:06	Split Factor	N/A
SampleType	Blank	Method Code	EU-047-NPW
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPfA	422-64-0	B210225-02211406	524	700	1530			L
ES	13C3-PFPfA		B210225-02211406				20-150%	97.1%	

Enthalpy Analytical

Job No.: 0225-819-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	MB_19032_PFAS	Prep Batch	EU19032
Sampling Site		Analyst	jogres
Enthalpy ID	MB_19032_PFAS	Instrument	Pippin
Matrix	aqueous	Sample Vol mL	250
Sampling Date		Extract Vol mL	0.4
Received Date		Split Factor	N/A
Prep Date	2025-02-28 09:17	Method Code	EU-047-NPW
AnalysisDate	2025-03-03 18:29		
SampleType	Blank		
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	P030325003	ND	0.254	0.640			U	
	PFPeA	2706-90-3	P030325003	ND	0.183	0.640			U	
	PFFhxA	307-24-4	P030325003	0.169	0.214	0.640			L	
	PFFHpA	375-85-9	P030325003	ND	0.224	0.640			U	
	PFOA	335-67-1	P030325003	ND	0.146	0.640			U	
	PFNA	375-95-1	P030325003	ND	0.145	0.640			U	
	PFDA	335-76-2	P030325003	ND	0.183	0.640			U	
	PFUnDA	2058-94-8	P030325003	ND	0.145	0.640			U	
	PFDoDA	307-55-1	P030325003	ND	0.260	0.640			U	
	PFTrDA	72629-94-8	P030325003	ND	0.212	0.640			U	
	PFTeDA	376-06-7	P030325003	ND	0.244	0.640			U	
	PFFhxDA	67905-19-5	P030325003	ND	0.340	0.640			U	
	Sulfonates	PFBS	375-73-5	P030325003	ND	0.340	0.640			U
		PFPeS	2706-91-4	P030325003	ND	0.131	0.603			U
		PFFhXS	355-46-4	P030325003	ND	0.494	0.586			U
		PFFHpS	375-92-8	P030325003	ND	0.310	0.610			U
PFOS		1763-23-1	P030325003	ND	0.338	0.593			U	
PFNS		68259-12-1	P030325003	ND	0.199	0.616			U	
PFDS		335-77-3	P030325003	ND	0.336	0.616			U	
4:2 FTS		757124-72-4	P030325003	ND	0.0830	0.600			U	
6:2 FTS		27619-97-2	P030325003	0.00160	0.302	0.610			L	
8:2 FTS		39108-34-4	P030325003	ND	0.143	0.613			U	
10:2 FTS	120226-60-0	P030325003	ND	0.490	0.640			U		
Sulfonamidos	FBSA	30334-69-1	P030325003	ND	0.304	0.640			U	
	N-EiFOSA	4151-50-2	P030325003	ND	0.396	0.640			U	
	N-EiFOSAA	2991-50-6	P030325003	ND	0.260	0.640			U	
	N-EiFOSE	1691-99-2	P030325003	ND	0.980	2.88			U	
	N-MeFOSA	31506-32-8	P030325003	ND	0.264	0.640			U	
	N-MeFOSAA	2355-31-9	P030325003	ND	0.180	0.640			U	
	N-MeFOSE	24448-09-7	P030325003	ND	0.608	2.88			U	
	PFOSA	754-91-6	P030325003	ND	0.0898	0.640			U	
	PFECAs	ADONA	919005-14-4	P030325003	ND	0.173	0.606			U
EVE Acid		69087-46-3	P030325003	ND	0.204	1.44			U	
HFPO-DA		13252-13-6	P030325003	ND	0.0678	0.640			U	
Hydro-EVE Acid		773804-62-9	P030325003	ND	0.210	0.640			U	
NFDHA		151772-58-6	P030325003	ND	0.135	0.640			U	
PEPA		267239-61-2	P030325003	ND	0.120	0.640			U	
PFECA-G		801212-59-9	P030325003	ND	0.0854	0.640			U	
PFMOAA		674-13-5	P030325003	ND	0.324	0.640			U	
PFMOBA		863090-89-5	P030325003	ND	1.07	1.44			U	
PFMOPrA		377-73-1	P030325003	ND	0.228	0.640			U	
PFO2HxA		39492-88-1	P030325003	ND	0.206	0.640			U	
PFO3OA		39492-89-2	P030325003	ND	0.294	0.640			U	
PFO4DA		39492-90-5	P030325003	0.241	0.506	3.20			L	
PFO5DA		39492-91-6	P030325003	0.169	0.512	3.20			L	
PMPA		13140-29-9	P030325003	ND	0.151	0.640			U	
R-EVE		2416366-22-6	P030325003	ND	1.06	1.44			U	
PFESAs		11CI-PF3OUdS	763051-92-9	P030325003	ND	0.302	0.603			U
		9CI-PF3ONS	756426-58-1	P030325003	ND	0.410	0.596			U
		Hydrolyzed PSDA	2416366-19-1	P030325003	0.154	0.426	0.640			L
	Nafion Byproduct 1 (PS Acid)	29311-67-9	P030325003	ND	0.342	0.640			U	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	P030325003	ND	0.530	0.640			U	
	NVHOS	1132933-86-8	P030325003	ND	0.0986	0.640			U	
	PFEESA	113507-82-7	P030325003	ND	0.192	0.640			U	
	R-PSDA	2416366-18-0	P030325003	ND	2.82	2.82			U	
	R-PSDCA	2416366-21-5	P030325003	ND	0.270	0.640			U	
	ES	MPFBA		P030325003				20-150%	93.4%	
M5PFPeA			P030325003				20-150%	87.0%		
M3PFBS			P030325003				20-150%	75.8%		
M2-4:2 FTS			P030325003				20-150%	133%		
M5PFFhxA			P030325003				20-150%	98.4%		
M3HFPO-DA			P030325003				20-150%	88.0%		
M4PFFHpA			P030325003				20-150%	103%		

Enthalpy Analytical

Job No.: 0225-819-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	MB_19032_PFAS	Prep Batch	EU19032
Sampling Site		Analyst	jogres
Enthalpy ID	MB_19032_PFAS	Instrument	Pippin
Matrix	aqueous	Sample Vol mL	250
Sampling Date		Extract Vol mL	0.4
Received Date		Split Factor	N/A
Prep Date	2025-02-28 09:17	Method Code	EU-047-NPW
AnalysisDate	2025-03-03 18:29		
SampleType	Blank		
Bottle ID	-		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M3PFHxS		P030325003				20-150%	105%	
M2-6:2 FTS		P030325003				20-150%	108%	
M8PFOA		P030325003				20-150%	94.8%	
M9PFNA		P030325003				20-150%	102%	
M8PFOS		P030325003				20-150%	96.7%	
M2-8:2 FTS		P030325003				20-150%	185%	Q
M8FOSA-I		P030325003				20-150%	55.1%	
M6PFDA		P030325003				20-150%	95.2%	
d3-N-MeFOSAA		P030325003				20-150%	99.4%	
d5-N-EtFOSAA		P030325003				20-150%	95.3%	
M7PFUdA		P030325003				20-150%	76.5%	
MPFDoA		P030325003				20-150%	55.4%	
M2PFTeDA		P030325003				20-150%	9.01%	Q
d3-N-MeFOSA		P030325003				10-200%	1.46%	Q
d5-N-EtFOSA		P030325003				10-200%	1.14%	Q
d7-N-MeFOSE		P030325003				10-200%	33.2%	
d9-N-EtFOSE		P030325003				10-200%	29.4%	

Enthalpy Analytical

Job No.: 0225-819-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	OPR_18976_PFAS		
Sampling Site			
Enthalpy ID	OPR_18976_PFAS	Prep Batch	eu18976
Matrix	aqueous	Analyst	ext-magennaef
Sampling Date		Instrument	Bumblebee
Received Date		Sample Vol mL	0.08
Prep Date	2025-02-18 12:15	Extract Vol mL	0.2
AnalysisDate	2025-02-21 14:18	Split Factor	N/A
SampleType	Control	Method Code	EU-047-NPW
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	B210225-02211418	28900	875	1910	40-150%	116%	
ES	13C3-PFPrA		B210225-02211418				20-150%	91.4%	

Enthalpy Analytical

Job No.: 0225-819-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Enthalpy ID	OPR_19032_PFAS	Prep Batch	EU19032	Sample Vol (mL)	250
Sample Name	OPR_19032_PFAS	Prep Date	2025-02-28 09:17	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2025-03-03 18:52	Split Factor	N/A
Sampling Date		Analyst	jogres	Method Code	EU-047-NPW
Received Date		Instrument	Pippin	Sample Type	Control
		Bottle ID	-		

	Compound	CAS	InjFileName	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	P030325004	17.2	0.254	0.640	47.9-144%	85.8%	
	PFPeA	2706-90-3	P030325004	16.9	0.183	0.640	41.7-159%	84.6%	
	PFHxA	307-24-4	P030325004	16.9	0.214	0.640	43.2-154%	84.5%	
	PFHpA	375-85-9	P030325004	15.7	0.224	0.640	42.1-155%	78.5%	
	PFOA	335-67-1	P030325004	16.5	0.146	0.640	51.1-148%	82.6%	
	PFNA	375-95-1	P030325004	16.7	0.145	0.640	51.6-153%	83.6%	
	PFDA	335-76-2	P030325004	17.7	0.183	0.640	44.5-156%	88.4%	
	PFUnDA	2058-94-8	P030325004	17.2	0.145	0.640	40.3-156%	86.0%	
	PFDoDA	307-55-1	P030325004	17.4	0.260	0.640	40.4-158%	86.9%	
	PFTtDA	72629-94-8	P030325004	46.6	0.212	0.640	42.2-201%	233%	Q
	PFTeDA	376-06-7	P030325004	18.0	0.244	0.640	43-162%	90.1%	
Sulfonates	PFBS	375-73-5	P030325004	14.8	0.340	0.640	42.7-155%	83.6%	
	PFPeS	2706-91-4	P030325004	15.7	0.131	0.603	40.3-152%	83.7%	
	PFHxS	355-46-4	P030325004	14.5	0.494	0.586	45-148%	79.5%	
	PFHpS	375-92-8	P030325004	16.8	0.310	0.610	39.8-166%	88.2%	
	PFOS	1763-23-1	P030325004	16.1	0.338	0.593	59.2-132%	86.8%	
	PFNS	68259-12-1	P030325004	15.1	0.199	0.616	38.1-153%	78.5%	
	PFDS	335-77-3	P030325004	12.3	0.336	0.616	28.6-148%	63.9%	
	4:2 FTS	757124-72-4	P030325004	17.1	0.0830	0.600	41.5-157%	91.4%	
	6:2 FTS	27619-97-2	P030325004	16.5	0.302	0.610	44.5-160%	86.6%	
8:2 FTS	39108-34-4	P030325004	17.3	0.143	0.613	39.4-166%	90.2%		
Sulfonamidos	N-EtFOSA	4151-50-2	P030325004	20.9	0.396	0.640	26.7-172%	105%	
	N-EtFOSAA	2991-50-6	P030325004	17.3	0.260	0.640	42.8-156%	86.5%	
	N-EtFOSE	1691-99-2	P030325004	85.5	0.980	2.88	38.9-161%	95.0%	
	N-MeFOSA	31506-32-8	P030325004	22.2	0.264	0.640	26.4-183%	111%	
	N-MeFOSAA	2355-31-9	P030325004	17.8	0.180	0.640	42-155%	89.2%	
	N-MeFOSE	24448-09-7	P030325004	83.6	0.608	2.88	37.6-155%	92.9%	
	PFOSA	754-91-6	P030325004	18.7	0.0898	0.640	39.1-158%	93.5%	
PFECAs	ADONA	919005-14-4	P030325004	16.3	0.173	0.606	32.2-151%	81.7%	
	HFPO-DA	13252-13-6	P030325004	18.6	0.0678	0.640	61.8-131%	93.1%	
PFESAs	11Cl-PF3OUdS	763051-92-9	P030325004	12.1	0.302	0.603	21.8-141%	60.3%	
	9Cl-PF3ONS	756426-58-1	P030325004	15.9	0.410	0.596	37.6-146%	79.3%	
ES	MPFBA		P030325004				20-150%	94.6%	
	M3PFPeA		P030325004				20-150%	90.7%	
	M3PFBS		P030325004				20-150%	78.8%	
	M2-4:2 FTS		P030325004				20-150%	127%	
	M5PFHxA		P030325004				20-150%	98.4%	
	M3HFPO-DA		P030325004				20-150%	89.1%	
	M4PFHpA		P030325004				20-150%	109%	
	M3PFHxS		P030325004				20-150%	105%	
	M2-6:2 FTS		P030325004				20-150%	106%	
	M8PFOA		P030325004				20-150%	100%	
	M9PFNA		P030325004				20-150%	107%	
	M8PFOS		P030325004				20-150%	97.0%	
	M2-8:2 FTS		P030325004				20-150%	190%	Q
	M8FOSA-I		P030325004				20-150%	69.6%	
	M6PFDA		P030325004				20-150%	100%	
	d3-N-MeFOSAA		P030325004				20-150%	101%	
	d5-N-EtFOSAA		P030325004				20-150%	97.4%	
	M7PFUdA		P030325004				20-150%	86.6%	
	MPFDoA		P030325004				20-150%	73.0%	
	M2PFTeDA		P030325004				20-150%	16.0%	Q
	d3-N-MeFOSA		P030325004				10-200%	1.70%	Q
	d5-N-EtFOSA		P030325004				10-200%	1.79%	Q
	d7-N-MeFOSE		P030325004				10-200%	43.2%	
d9-N-EtFOSE		P030325004				10-200%	38.0%		

Sample Custody



0225-819

Chain of Custody Record

Enthalpy Ultratrace Job#: _____ COC Page 1 of 1

Special Handling:

- Standard Turn Around Time
- Rush Turn Around Time -- Date Needed _____
- All Fast TATs Subject to Approval by Enthalpy Analytical, Inc.
- All Samples Disposed of After 6 months Unless Otherwise Instructed.

Enthalpy Analytical-Wilmington, NC has added enhancements to standard methods to improve accuracy, precision and permit an assessment of laboratory performance in the context of your specific data needs. For more information email Cindy.James@enthalpy.com.

Client Name: BRUNSWICK COUNTY UTILITIES
 Project Manager: GLENN WALKER
 Report To: SAME

Project Number: _____
 Site Name: NORTHWEST WATER PLANT
 Location: LELAND N.C.

PO#: _____
 Telephone#: _____
 Email: _____

This Chain of Custody is applicable to Non-Air samples. Standard TAT differ per analysis and are provided by request.

Client Special Instructions:

Matrix: GW-Groundwater, WW-Wastewater, NW-Non-Potable Water, DW-Drinking Water, S-Soil, SL-Sludge, BT-Biological Tissue, O-Other
 Type: G=Grab C=Composite Q=Quality Control

Sample ID	Date	Time	Sample Volume	Type	Matrix	Sample Containers				Analyses:							Notes:		
						# of Bottles	# of Jars	# of Bags	# Other	Method 1613	Method 8290	Method 1668A/B/C PCE	PFAS by LC/MS/MS	PAHs by HRGC/HRMS	Sample on Hold	Method 23		ALL PFAS	
021425-SO1	2/14/2025	1305PM	250 ml	G	NW	2												X	Please Add PFPrA and
021425-EO1	2/14/2025	1305PM	250 ml	G	DW	2												X	PFHpA To The Testing.
																			Mark Hager Knows About
																			This if you Have Questions

Relinquished By	Date	Received By	Date	Time	Sample Temperature Upon Receipt
	BILLY BENTON 2/115/25		2/11/25	14:40	<input checked="" type="checkbox"/> Iced <input type="checkbox"/> Ambient °C <u>1.3</u>
					<input type="checkbox"/> Iced <input type="checkbox"/> Ambient °C _____
					<input type="checkbox"/> Iced <input type="checkbox"/> Ambient °C _____

JOB ID: 0225-819 Date / Time: 2/14/25 14:40 Initials: M.A.C
 OR
 Client: Brunswick County Public Utilities

Cooler 1 of 1

Temp °C: 1.3 Thermometer ID: T16

- Received via
 FedEx
 UPS
 DHL
 USPS
 Courier
 Other

Check one
 On ice:
 Melted ice:
 Ambient:

Check one
 in a Box:
 in a Cooler:
 Cooler in Box:

	Yes	No
Cooler seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Sample seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Good condition:	<input checked="" type="checkbox"/>	<input type="checkbox"/>

Comment:

Temp °C: Thermometer ID: Cooler of

Received via
 FedEx
 UPS
 DHL
 USPS
 Courier
 Other

Check one
 On ice:
 Melted ice:
 Ambient:

Check one
 in a Box:
 in a Cooler:
 Cooler in Box:

	Yes	No
Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>
Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>
Good condition:	<input type="checkbox"/>	<input type="checkbox"/>

Comment:

Temp °C: Thermometer ID: Cooler of

Received via
 FedEx
 UPS
 DHL
 USPS
 Courier
 Other

Check one
 On ice:
 Melted ice:
 Ambient:

Check one
 in a Box:
 in a Cooler:
 Cooler in Box:

	Yes	No
Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>
Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>
Good condition:	<input type="checkbox"/>	<input type="checkbox"/>

Comment: