

October 19, 2022

Glenn Walker
Brunswick County Water Systems
PO Box 249
Bolivia, NC 28422

RE: Project: 1,4-Dx-522(WEEKLY)
Pace Project No.: 35751475

Dear Glenn Walker:

Enclosed are the analytical results for sample(s) received by the laboratory on October 07, 2022. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Ormond Beach

If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Bill White
bill.white@pacelabs.com
(386) 672-5668
Project Manager

Enclosures

cc: Billy Benton, BRUNSWICK COUNTY PUBLIC UTILITIES



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

CERTIFICATIONS

Project: 1,4-Dx-522(WEEKLY)

Pace Project No.: 35751475

Pace Analytical Services Ormond Beach

8 East Tower Circle, Ormond Beach, FL 32174

Alaska DEC- CS/UST/LUST

Alabama Certification #: 41320

Colorado Certification: FL NELAC Reciprocity

Connecticut Certification #: PH-0216

Delaware Certification: FL NELAC Reciprocity

Florida Certification #: E83079

Georgia Certification #: 955

Guam Certification: FL NELAC Reciprocity

Hawaii Certification: FL NELAC Reciprocity

Illinois Certification #: 200068

Indiana Certification: FL NELAC Reciprocity

Kansas Certification #: E-10383

Kentucky Certification #: 90050

Louisiana Certification #: FL NELAC Reciprocity

Louisiana Environmental Certificate #: 05007

Maine Certification #: FL01264

Maryland Certification: #346

Massachusetts Certification #: M-FL1264

Michigan Certification #: 9911

Mississippi Certification: FL NELAC Reciprocity

Missouri Certification #: 236

Montana Certification #: Cert 0074

Nebraska Certification: NE-OS-28-14

New Hampshire Certification #: 2958

New Jersey Certification #: FL022

New York Certification #: 11608

North Carolina Environmental Certificate #: 667

North Carolina Certification #: 12710

North Dakota Certification #: R-216

Ohio DEP 87780

Oklahoma Certification #: D9947

Pennsylvania Certification #: 68-00547

Puerto Rico Certification #: FL01264

South Carolina Certification: #96042001

Tennessee Certification #: TN02974

Texas Certification: FL NELAC Reciprocity

US Virgin Islands Certification: FL NELAC Reciprocity

Virginia Environmental Certification #: 460165

West Virginia Certification #: 9962C

Wisconsin Certification #: 399079670

Wyoming (EPA Region 8): FL NELAC Reciprocity

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE SUMMARY

Project: 1,4-Dx-522(WEEKLY)

Pace Project No.: 35751475

Lab ID	Sample ID	Matrix	Date Collected	Date Received
35751475001	100622S01	Water	10/06/22 07:15	10/07/22 10:40
35751475002	100622E01	Drinking Water	10/06/22 07:15	10/07/22 10:40

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE ANALYTE COUNT

Project: 1,4-Dx-522(WEEKLY)

Pace Project No.: 35751475

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
35751475001	100622S01	EPA 522	BMH	2	PASI-O
35751475002	100622E01	EPA 522	BMH	2	PASI-O

PASI-O = Pace Analytical Services - Ormond Beach

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: 1,4-Dx-522(WEEKLY)

Pace Project No.: 35751475

Sample: 100622S01									
Lab ID: 35751475001									
Collected: 10/06/22 07:15 Received: 10/07/22 10:40 Matrix: Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
522 MSS 1,4 Dioxane									
Analytical Method: EPA 522 Preparation Method: EPA 522									
Pace Analytical Services - Ormond Beach									
1,4-Dioxane (p-Dioxane)	0.97	ug/L	0.19	0.12	1	10/17/22 10:45	10/18/22 20:09	123-91-1	
Surrogates									
1,4-Dioxane-d8 (S)	82	%	70-130		1	10/17/22 10:45	10/18/22 20:09		

Sample: 100622E01									
Lab ID: 35751475002									
Collected: 10/06/22 07:15 Received: 10/07/22 10:40 Matrix: Drinking Water									
Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
522 MSS 1,4 Dioxane									
Analytical Method: EPA 522 Preparation Method: EPA 522									
Pace Analytical Services - Ormond Beach									
1,4-Dioxane (p-Dioxane)	0.83	ug/L	0.20	0.12	1	10/17/22 10:45	10/18/22 20:27	123-91-1	
Surrogates									
1,4-Dioxane-d8 (S)	79	%	70-130		1	10/17/22 10:45	10/18/22 20:27		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: 1,4-Dx-522(WEEKLY)
Pace Project No.: 35751475

QC Batch: 864066	Analysis Method: EPA 522
QC Batch Method: EPA 522	Analysis Description: 522 MSS 1,4 Dioxane
	Laboratory: Pace Analytical Services - Ormond Beach

Associated Lab Samples: 35751475001, 35751475002

METHOD BLANK: 4755490 Matrix: Water
Associated Lab Samples: 35751475001, 35751475002

Parameter	Units	Blank Result	Reporting Limit	MDL	Analyzed	Qualifiers
1,4-Dioxane (p-Dioxane)	ug/L	0.12 U	0.20	0.12	10/18/22 14:47	
1,4-Dioxane-d8 (S)	%	86	70-130		10/18/22 14:47	

LABORATORY CONTROL SAMPLE: 4755491

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,4-Dioxane (p-Dioxane)	ug/L	20	20.3	102	70-130	
1,4-Dioxane-d8 (S)	%			92	70-130	

LABORATORY CONTROL SAMPLE: 4755492

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
1,4-Dioxane (p-Dioxane)	ug/L	0.2	0.14 I	71	50-150	
1,4-Dioxane-d8 (S)	%			92	70-130	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 4755493 4755494

Parameter	Units	MS		MSD		MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
		60411116005 Result	Spike Conc.	Spike Conc.	Result						
1,4-Dioxane (p-Dioxane)	ug/L	ND	19.3	19.4	19.8	102	102	70-130	1	20	
1,4-Dioxane-d8 (S)	%					94	95	70-130			

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALIFIERS

Project: 1,4-Dx-522(WEEKLY)

Pace Project No.: 35751475

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

ANALYTE QUALIFIERS

I The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.

U Compound was analyzed for but not detected.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: 1,4-Dx-522(WEEKLY)
Pace Project No.: 35751475

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
35751475001	100622S01	EPA 522	864066	EPA 522	864344
35751475002	100622E01	EPA 522	864066	EPA 522	864344

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

WO#: 35751475



CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

Required Client Information:
 Company: Brunswick County Water Systems
 Address: PO Box 249
 Bolivia, NC 28422
 Email To: Glenn Walker
 Phone: 910-371-3490 Fax:
 Requested Due Date: W007

Required Project Information:
 Report To: Glenn Walker
 Copy To:
 Purchase Order #: 1,4-Dx-522 (Weekly)
 Project Name: Lisa Harvey
 Project #: 9551-1 (SO1) , -2 (EO1)

Invoice Information:
 Attention: Accounts Payable
 Company Name: See Section A
 Address:
 Pace Quote:
 Pace Project Manager: Lisa Harvey
 Pace Profile #: 9551-1 (SO1) , -2 (EO1)

Regulatory Agency: NC
State / Location: NC

ITEM #	MATRIX	CODE	MATRIX CODE (see valid codes to left)	SAMPLE TYPE (G=GRAB C=COMP)	COLLECTED		DATE	TIME	DATE	TIME	ACCEPTED BY / AFFILIATION	DATE	TIME	SAMPLE CONDITIONS
					START	END								
1	Drinking Water	DW	WT G	WT G	10/06/22	715	10/06/22	715	1	1330	Phil McCulloch	10/06/2022	1330	line 1
2	Waste Water	WW	DW G	DW G	10/06/22	715	10/06/22	715	1	1330	Phil McCulloch	10/06/2022	1330	line 2
3	Product	P												
4	Soil/Solid	SL												
5	Oil	OL												
6	Wipe	WP												
	Air	AR												
	Other	OT												
	Tissue	TS												
Requested Analysis Filtered (Y/N)														
Preservatives														
H2SO4														
HNO3														
HCl														
NaOH														
Na2SO3														
Methanol														
Sodium sulfite and sodium bisulfate														
Analyses Test														
522 - 1,4 - Dioxane														
522 - 1,4 - Dioxane														
Residual Chlorine (Y/N)														
line 1														
line 2														

SAMPLER NAME AND SIGNATURE

PRINT Name of SAMPLER: Phil McCulloch
 SIGNATURE of SAMPLER: *Phil McCulloch*
 DATE Signed: 10/06/22

SO1 is Raw Water (WT)
 EO1 is Potable DWTR (DW)



WO# : 35751475
 PM: **WBW** Due Date: **10/19/22**
 CLIENT: **BRUNCOWS**

Project #
 Project Manager:
 Client:

Date and Initials of person:
 Examining contents:
 Label:
 Deliver:
 pH:

Thermometer Used: T-392 Date: 10/1/22 Time: 1049 Initials: BUJ

State of Origin: _____ For WV projects, all containers verified to ≤6 °C

Cooler #1 Temp. °C 1.0 (Visual) 10.2 (Correction Factor) 1.2 (Actual) Samples on ice, cooling process has begun

Cooler #2 Temp. °C _____ (Visual) _____ (Correction Factor) _____ (Actual) Samples on ice, cooling process has begun

Cooler #3 Temp. °C _____ (Visual) _____ (Correction Factor) _____ (Actual) Samples on ice, cooling process has begun

Cooler #4 Temp. °C _____ (Visual) _____ (Correction Factor) _____ (Actual) Samples on ice, cooling process has begun

Cooler #5 Temp. °C _____ (Visual) _____ (Correction Factor) _____ (Actual) Samples on ice, cooling process has begun

Cooler #6 Temp. °C _____ (Visual) _____ (Correction Factor) _____ (Actual) Samples on ice, cooling process has begun

Recheck for OOT °C _____ (Visual) _____ (Correction Factor) _____ (Actual) Time: _____ Initials: _____

Courier: Fed Ex UPS USPS Client Commercial Pace Other _____

Shipping Method: First Overnight Priority Overnight Standard Overnight Ground International Priority

Billing: Recipient Sender Third Party Credit Card Unknown

Tracking # 2788 3438 0248

Custody Seal on Cooler/Box Present: Yes No Seals intact: Yes No Ice: Wet Blue Melted None

Packing Material: Bubble Wrap Bubble Bags None Other _____

Samples shorted to lab (If Yes, complete) Shorted Date: _____ Shorted Time: _____ Qty: _____

	Yes	No	N/A	Comments:
Chain of Custody Present	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Chain of Custody Filled Out	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Relinquished Signature & Sampler Name COC	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Samples Arrived within Hold Time	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Rush TAT requested on COC	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Sufficient Volume	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Correct Containers Used	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Containers Intact	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Sample Labels match COC (sample IDs & date/time of collection)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
All containers needing acid/base preservation have been checked.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
All Containers needing preservation are found to be in compliance with EPA recommendation:	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Preservation Information: Preservative: _____ Lot #/Trace #: _____ Date: _____ Time: _____ Initials: _____
Exceptions: Vials, Microbiology, O&G, PFAS	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Headspace in VOA Vials? (>6mm):	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Trip Blank Present:	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	

Comments/ Resolution (use back for additional comments):

