

Brunswick County Public Utilities - NC

PO Box 249
Bolivia, NC 28422-0249

LELAND N.C.

Client Project# NORTHWEST WATER PLANT
Samples Received: 8/1/2025

Analytical Report 0825-736

PFAS by Isotope Dilution (non-potable water)

Report Issue Date: 8/30/2025

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This analytical report was prepared in Portable Document Format (.PDF) and contains 28 pages. This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:



Alexandra Mejia, Quality Assurance Associate I



Enthalpy Analytical, LLC – Wilmington
Amanda Valois, Project Manager Assistant
amvalois@montrose-env.com / www.enthalpy.com
O: (910) 212-5858
2714 Exchange Drive, Wilmington, NC 28405

Table of Contents

Case Narrative	3
General Reporting Notes	6
Results	12
Summary of Results	13
Detailed Results	14
QC Data	20
Blanks	21
Controls	24
Sample Custody	26
Chain of Custody	27

Narrative Summary



Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0825-736-1
Client ID.	NORTHWEST WATER PLANT Site: LELAND N.C.

1. Custody

Jayson-Shane Santos received the samples at 5.6 °C after being relinquished by Brunswick County Public Utilities - NC.

The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC.

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix	Received
0825-736-001-1	080125-SO1	aqueous	2025-08-01
0825-736-001-2	080125-SO1	aqueous	2025-08-01
0825-736-002-1	080125-EO1	aqueous	2025-08-01
0825-736-002-2	080125-EO1	aqueous	2025-08-01

2. Methods and Analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU047	Brunswick List	ENVI-Carb

3. Analysis

The samples were analyzed using LC/MS/MS instruments Bumblebee and Pippin.

Polar compound PFPrA in the samples, including the method blank (MB) and Ongoing Precision Recovery (OPR) samples, was analyzed by direct inject calibration.

The samples were analyzed using more than one batch preparation and analytical sequence to include all analytes of interest.

4. Calibration

In the initial calibration, the reported analytes exhibited R^2 of ≥ 0.99 . The reported analytes in the calibration standards, Initial Calibration Verification (ICV) and continuing calibration (concal) met the accuracy criterion for native analytes.

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0825-736-1
Client ID.	NORTHWEST WATER PLANT Site: LELAND N.C.

The Standards that did not were:

- SID BM67 (PFO4DA, PFO5DA)
- SID BM68 (PFO4DA, PFO5DA)

Analyte(s) that exceed method control limits in the concals were not detected >LOQ in the samples. The data is reported without adverse impact.

5. QC Notes

The QC sample analyses passed all method criteria.

PFAS by Isotope Dilution (non-potable water) samples were extracted within 28 days, and extracts analyzed within 28 days.

6. Reporting Notes

The results presented in this report are representative of the samples as provided to the laboratory.

This report provides all results including detections below LOD following client instruction.

Some labeled extraction standards (ES) in the analyses recovered outside method control limits for ES recovery, as denoted by the "Q" qualifier. The target analytes are quantified based on their ratio to their labeled standard analogs. As a result, low or high labeled standard recovery do not cause any change to ratios or contribute any additional error in the measurement of the target analytes. When detected at a signal-to-noise above 10:1 the ES peak area is used to quantify its respective target analyte using accepted isotope dilution principles. The data is reported without adverse impact.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2016 TNI Standard under certificate number 05075.

General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, LLC - Wilmington, NC data reports, unless specifically noted otherwise.

General Data Qualifiers

- Ac - Alternate calculation flag indicates the es recovery was calculated using the opening concal when either of the following situations is encountered in the data processing software: the ES recovery is over 400% or the JS is not detected.
- B – The analyte was found in the method blank, at a concentration that was at least 10% of the amount in the sample.
- Cxx – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group ('xx') are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve). For HRMS data, this condition does not imply additional measurement uncertainty. For LC-MS/MS data, these values should be considered as having measurement uncertainty higher than values within the calibration range.
- EDL – Estimated Detection Level: The EDL is unique to isotope dilution methods and reflects the conditions of analysis at the time of analysis, including the equipment used. Where the MDL is a static value, the EDL is a dynamic value.
- EMPC – Estimated Maximum Possible Concentration: EMPC is specific to Dioxin/Furan tests to indicate the determined ion-abundance ratio was outside the allowed theoretical range (usually due to being near the detection limit, although it can very rarely be caused by a co-eluting interference). The EMPC concentration is adjusted to reflect the value at the theoretical ion-abundance ratio.
- I/IR – The ion ratio between the primary and secondary ions was observed to be outside the method criteria. The analyte concentration may be inaccurate due to interference.
- J – The analyte has a concentration below the minimum calibration level (LOQ value) but greater than the LOD. These values should be considered as having measurement uncertainty higher than values within the calibration range
- L - For reports containing PFAS analytes only, this flag indicates that an analyte has a concentration below the Minimum Detection Limit (MDL) . The reported concentration is not recommended for regulatory use as the analyte signal may have a signal-to-noise ratio less than the criteria deemed necessary to be considered a detected analyte.
- LOD – Limit of Detection: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOD. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the minimum detection limit (MDL). The LOD is adjusted for sample weight or volume.



General Reporting Notes – Data Qualifiers

- LOQ – Limit of Quantitation: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOQ. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the reporting limit (RL). The LOQ is adjusted for sample weight or volume.
- <LOD() – Analyte was not found at a concentration high enough to be reported as detected. It is reported as less than the LOD, and the LOD is given in the parentheses.
- <LOQ() – Analyte was not found at a concentration high enough to be reported as above the QSM-defined LOQ or TNI defined Reporting Limit. It is reported as less than the LOQ, and the LOQ is given in the parentheses.
- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable due to issues observed in sample preparation or analysis.
- PR – The associated congener(s) is(are) poorly resolved.
- QI – Indicates the presence of a quantitative interference.
- RL – Reporting Limit. Lowest reportable value. The level is higher than the MDL.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected.
- V / Q – The labeled standard recovery is not within method control limits.
- X – Indicates the result is from re-injection/repeat/second-column analysis.

Lab Identifiers/ Data Attributes

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Dilution Data. Result was obtained from the analysis of a dilution. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.



General Reporting Notes – Data Qualifiers

- R – Indicates a re-extraction of the sample.
- RJ – Indicates a reinjection of the sample extract.
- S – Indicates a sample split. The number that follows the “S” indicates the split factor.
- SAT – Indicates an analyte saturated the detector.

PFAS Compound Acronym List			Methods					
Acronym	CAS #	Compound Name	SOP EU047	EPA 1633 (B-24)	EPA 1633X	EPA 537.1	EPA 533	EPA 8327*
Target Analytes								
PFBA	375-22-4	Perfluorobutanoic Acid	X	X	X		X	X
PFPeA	2706-90-3	Perfluoropentanoic Acid	X	X	X		X	X
PFHxA	307-24-4	Perfluorohexanoic Acid	X	X	X	X	X	X
PFHpA	375-85-9	Perfluoroheptanoic Acid	X	X	X	X	X	X
PFOA	335-67-1	Perfluorooctanoic Acid	X	X	X	X	X	X
PFNA	375-95-1	Perfluorononanoic Acid	X	X	X	X	X	X
PFDA	335-76-2	Perfluorodecanoic acid	X	X	X	X	X	X
PFUnA (PFUnDA)	2058-94-8	Perfluoroundecanoic acid	X	X	X	X	X	X
PFDoA (PFDoDA)	307-55-1	Perfluorododecanoic acid	X	X	X	X		X
PFTrDA (PFTriA, PFTrDA)	72629-94-8	Perfluorotridecanoic acid	X	X	X	X		X
PFTeDA (PFTA, PFTreA)	376-06-7	Perfluorotetradecanoic acid	X	X	X	X		X
PFBS	375-73-5	Perfluorobutane sulfonic acid	X	X	X	X	X	X
PFPeS	2706-91-4	Perfluoropentane sulfonic acid	X	X	X		X	X
PFHxS	355-46-4	Perfluorohexane sulfonic acid	X	X	X	X	X	X
PFHpS	375-92-8	Perfluoroheptane sulfonic acid	X	X	X		X	X
PFOS	1763-23-1	Perfluorooctane sulfonic acid	X	X	X	X	X	X
PFNS	68259-12-1	Perfluorononane sulfonic acid	X	X	X			X
PFDS	335-77-3	Perfluorodecane sulfonic acid	X	X	X			X
4:2 FTS	757124-72-4	4:2 fluorotelomer sulfonic acid	X	X	X		X	X
6:2 FTS	27619-97-2	6:2 fluorotelomer sulfonic acid	X	X	X		X	X
8:2 FTS	39108-34-4	8:2 fluorotelomer sulfonic acid	X	X	X		X	X
10:2 FTS	120226-60-0	Fluorotelomer sulfonate 10:2						X
FHxSA	41997-13-1	Perfluorohexanesulfonamide			X			X
PFOSA (FOSA)	754-91-6	Perfluorooctane sulfonamide	X	X	X			X
N-MeFOSAA	2355-31-9	N-methyl perfluorooctane sulfonamido acetic acid	X	X	X	X		X
N-MeFOSA	31506-32-8	N-methylperfluoro-1-octanesulfonamide	X	X	X			X
N-MeFOSE	24448-09-7	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol	X	X	X			X
N-EtFOSAA	2991-50-6	N-ethyl perfluorooctane sulfonamido acetic acid	X	X	X	X		X
N-EtFOSA	4151-50-2	N-ethylperfluoro-1-octanesulfonamide	X	X	X			X
N-EtFOSE	1691-99-2	2-(N-Ethylperfluoro-1-octanesulfonamido)-ethanol	X	X	X			X
HFPO-DA	13252-13-6	Hexafluoropropyleneoxide dimer acid (GenX)	X	X	X	X	X	X
11Cl-PF3OUds	763051-92-9	11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	X	X	X	X	X	X
9Cl-PF3ONS	756426-58-1	9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	X	X	X	X	X	X
ADONA	919005-14-4	4,8-dioxa-3H-perfluorononanoic acid	X	X	X	X	X	X
PFESA	113507-82-7	Perfluoro(2-ethoxyethane)sulfonic acid		X	X		X	X
PFMOBA (PFMBA)	863090-89-5	Perfluoro-4-methoxybutanoic acid		X	X		X	X
NFDHA	151772-58-6	Nonafluoro-3,6-dioxaheptanoic acid		X	X		X	X
PFMOPrA (PFMPA)	377-73-1	Perfluoro-3-methoxypropanoic acid		X	X		X	X
PFPrA	422-64-0	Perfluoropropionic acid, 2,2,3,3-Pentafluoropropionic acid			X			X
PFPrS (PFPS)	423-41-6	Perfluoropropanesulfonic acid			X			X



PFAS Compound Acronym List			Methods					
Acronym	CAS #	Compound Name	SOP EU047	EPA 1633 (B-24)	EPA 1633X	EPA 537.1	EPA 533	EPA 8327*
PFMOAA	674-13-5	Perfluoro-2-methoxyacetic acid;			X			X
PFO2HxA	39492-88-1	Perfluoro (3,5-dioxahexanoic) acid			X			X
PFO3OA	39492-89-2	Perfluoro (3,5,7-trioxaoctanoic) acid			X			X
PFO4DA	39492-90-5	Perfluoro (3,5,7,9-tetraoxadecanoic) acid			X			X
PFO5DA	39492-91-6	Perfluoro(3,5,7,9,11-pentaoxadodecanoic) acid			X			X
Nafion Byproduct 1 (PS Acid)	29311-67-9	1,1,2,2-tetrafluoro-2-[1,1,1,2,3,3-hexafluoro-3-(1,2,2-trifluoroethenoxy)propan-2-yl]oxyethanesulfonic acid			X			X
Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	Perfluoro-2-[[perfluoro-3-(perfluoroethoxy)-2-propanyl]oxy]ethanesulfonic acid (Hydro-PS Acid)			X			X
PEPA	267239-61-2	Perfluoro-2-ethoxypropanoic acid			X			X
PMPA	13140-29-9	Perfluoro-2-methoxypropanoic acid			X			X
PFECA-G, (PFPE-1)	801212-59-9	4-(Heptafluoroisopropoxy)hexafluorobutanoic acid, Perfluoro-4-isopropoxybutanoic acid			X			X
PFHxDA	67905-19-5	Perfluorohexadecanoic acid			X			
R-PSDA (Nafion Byproduct 4)	2416366-18-0	Perfluoro-4-(2-sulfoethoxy)pentanoic acid; 2,2,3,3,4,5,5-Octafluoro-4-(1,1,2,2-tetrafluoro-2-sulfoethoxy)pentanoic acid			X			X
Hydrolyzed PSDA (Nafion Byproduct 5)	2416366-19-1	2-fluoro-2-[1,1,2,3,3,3-hexafluoro-2-(1,1,2,2-tetrafluoro-2-sulfoethoxy)propoxy]-acetic acid			X			X
R-PSDCA (Nafion Byproduct 6)	2416366-21-5	1,1,2,2-tetrafluoro-2-[1,2,2,3,3-pentafluoro-1-(trifluoromethyl)propoxy] ethanesulfonic acid			X			X
EVE Acid	69087-46-3	2,2,3,3-tetrafluoro-3-((1,1,1,2,3,3-hexafluoro-3-[(1,2,2-trifluoroethenyl)oxy]propan-2-yl)oxy)propionic acid			X			X
FBSA	30334-69-1	Perfluorobutylsulfonamide			X			X
MeFBSA	68298-12-4	1-Butanesulfonamide; (N-(Methyl)nonafluorobutanesulfonamide); 1,1,2,2,3,3,4,4,4-nonafluoro-N-methyl-1-Butanesulfonamide			X			X
Hydro-EVE Acid	773804-62-9	2,2,3,3-Tetrafluoro-3-[[1,1,1,2,3,3-hexafluoro-3-(1,2,2,2-tetrafluoroethoxy)propan-2-yl]oxy}propanoic acid			X			X
R-EVE Acid	2416366-22-6	4-(2-carboxy-1,1,2,2-tetrafluoroethoxy)-2,2,3,3,4,5,5,5-octafluoro-pentanoic acid			X			X
NVHOS	1132933-86-8	Perfluoroethoxysulfonic acid; 1,1,2,2-Tetrafluoro-2-(1,2,2,2-tetrafluoroethoxy)ethane-1-sulfonic acid			X			X

PFAS Compound Acronym List			Methods					
Acronym	CAS #	Compound Name	SOP EU047	EPA 1633 (B-24)	EPA 1633X	EPA 537.1	EPA 533	EPA 8327*
PFDoS	79780-39-5	Perfluorododecane sulfonic acid		X	X			X
PFOA	16517-11-6	Perfluorooctadecanoic acid			X			
3:3 FTCA	356-02-5	2H,2H,3H,3H-Perfluorohexanoic acid		X	X			X
5:3 FTCA	914637-49-3	2H,2H,3H,3H-Perfluorooctanoic acid		X	X			X
7:3 FTCA	812-70-4	2H,2H,3H,3H-Perfluorodecanoic acid		X	X			X
N-AP-FHxSA	50598-28-2	N-(3-(Dimethylamino)propyl)tridecafluoro-1-hexanesulfonamide			X			X
N-CMAmP-6:2 FOSA	34455-29-3	N-(Carboxymethyl)-N,N-dimethyl-3-(((3,3,4,4,5,5,6,6,7,7,8,8,8-tridecafluorooctyl)sulfonyl)amino)1-propanaminium			X			X
BPAF	1478-61-1	Bisphenol AF			X			X
HQ-115	90076-65-6	Bis(trifluoromethane)sulfonimide lithium salt			X			X

* Accreditation pending

Results

Enthalpy Analytical

Job No.: 0825-736-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Summary

	Compound	CAS	080125-SO1 ng/L	080125-SO1 ng/L	080125-EO1 ng/L	080125-EO1 ng/L
Acids	PFPtA	422-64-0	421 L		419 L	
	PFBA	375-22-4		5.20		4.29
	PFPeA	2706-90-3		6.60		5.80
	PFHxA	307-24-4		5.42		5.06
	PFHpA	375-85-9		2.89		2.75
	PFOA	335-67-1		6.80		6.36
	PFNA	375-95-1		1.11		1.00
	PFDA	335-76-2		0.741		0.702
	PFUnDA	2058-94-8		0.166 J		0.118 J
	PFDoDA	307-55-1		ND U		ND U
	PFTtDA	72629-94-8		ND U		ND U
	PFTeDA	376-06-7		ND U		ND U
	PFHxDA	67905-19-5		ND U		ND U
	Sulfonates	PFBS	375-73-5		3.44	
PFPeS		2706-91-4		0.532		0.501
PFHxS		355-46-4		3.62		3.17
PFHpS		375-92-8		0.107 L		0.0967 L
PFOS		1763-23-1		14.2		13.2
PFNS		68259-12-1		ND U		ND U
PFDS		335-77-3		ND U		ND U
4:2 FTS		757124-72-4		ND U		ND U
6:2 FTS		27619-97-2		ND U		ND U
8:2 FTS		39108-34-4		ND U		0.00322 L
10:2 FTS		120226-60-0		ND U		ND U
Sulfonamidos		FBSA	30334-69-1		0.347 J	
	N-EtFOSA	4151-50-2		ND U		ND U
	N-EtFOSAA	2991-50-6		0.0829 L		0.0441 L
	N-EtFOSE	1691-99-2		ND U		ND U
	N-MeFOSA	31506-32-8		ND U		ND U
	N-MeFOSAA	2355-31-9		0.00101 L		ND U
	N-MeFOSE	24448-09-7		ND U		ND U
	PFOSA	754-91-6		0.0456 L		ND U
PFECAs	ADONA	919005-14-4		ND U		ND U
	EVE Acid	69087-46-3		ND U		ND U
	HFPO-DA	13252-13-6		1.61		1.58
	Hydro-EVE Acid	773804-62-9		ND U		ND U
	NFDHA	151772-58-6		ND U		ND U
	PEPA	267239-61-2		1.10		1.04
	PFECA-G	801212-59-9		ND U		ND U
	PFMOAA	674-13-5		13.0		10.9
	PFMOBA	863090-89-5		ND U		ND U
	PFMOPrA	377-73-1		ND U		ND U
	PFO2HxA	39492-88-1		3.54		3.31
	PFO3OA	39492-89-2		1.31		1.13
	PFO4DA	39492-90-5		ND U		ND U
	PFO5DA	39492-91-6		ND U		ND U
	PMPA	13140-29-9		2.22		2.27
	R-EVE	2416366-22-6		2.28		1.99
	PFESAs	11Cl-PF3OUdS	763051-92-9		ND U	
9Cl-PF3ONS		756426-58-1		ND U		ND U
Hydrolyzed PSDA		2416366-19-1		1.26		1.44
Nafion Byproduct 1 (PS Acid)		29311-67-9		ND U		ND U
Nafion Byproduct 2 (Hydro-PS Acid)		749836-20-2		0.204 L		0.233 L
NVHOS		1132933-86-8		1.36		1.64
PFEESA		113507-82-7		ND U		ND U
R-PSDA		2416366-18-0		3.74		4.34
R-PSDCA		2416366-21-5		ND U		ND U

Enthalpy Analytical

Job No.: 0825-736-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	080125-SO1		
Sampling Site			
Enthalpy ID	0825-736-001-1	Prep Batch	EU118522
Matrix	aqueous	Analyst	jogres
Sampling Date	2025-08-01 13:35	Instrument	Bumblebee
Received Date	2025-08-01	Sample Vol mL	0.1
Prep Date	2025-08-05 12:30	Extract Vol mL	0.2
AnalysisDate	2025-08-05 22:25	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	A		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	8050825-08052225	421	700	1530			L
ES	13C3-PFPrA		8050825-08052225				20-150%	86.4%	

Enthalpy Analytical

Job No.: 0825-736-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name 080125-SO1
 Sampling Site
 Enthalpy ID 0825-736-001-2 Prep Batch EU118528
 Matrix aqueous Analyst jogres
 Sampling Date 2025-08-01 13:35 Instrument Pippin
 Received Date 2025-08-01 Sample Vol mL 302.32
 Prep Date 2025-08-06 09:05 Extract Vol mL 0.4
 AnalysisDate 2025-08-08 01:12 Split Factor N/A
 SampleType Sample Method Code EU-047-NPW
 Bottle ID A

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	P070825026	5.20	0.210	0.529				
	PFPeA	2706-90-3	P070825026	6.60	0.151	0.529				
	PFHxA	307-24-4	P070825026	5.42	0.177	0.529				
	PFFHpA	375-85-9	P070825026	2.89	0.185	0.529				
	PFOA	335-67-1	P070825026	6.80	0.121	0.529				
	PFNA	375-95-1	P070825026	1.11	0.120	0.529				
	PFDA	335-76-2	P070825026	0.741	0.151	0.529				
	PFUnDA	2058-94-8	P070825026	0.166	0.120	0.529			J	
	PFDoDA	307-55-1	P070825026	ND	0.215	0.529			U	
	PFTrDA	72629-94-8	P070825026	ND	0.175	0.529			U	
	PFTeDA	376-06-7	P070825026	ND	0.202	0.529			U	
	PFFhxDA	67905-19-5	P070825026	ND	0.281	0.529			U	
	Sulfonates	PFBS	375-73-5	P070825026	3.44	0.281	0.529			
PFPeS		2706-91-4	P070825026	0.532	0.109	0.499				
PFFhXS		355-46-4	P070825026	3.62	0.409	0.485				
PFFHpS		375-92-8	P070825026	0.107	0.256	0.504			L	
PFOS		1763-23-1	P070825026	14.2	0.280	0.490				
PFNS		68259-12-1	P070825026	ND	0.164	0.510			U	
PFDS		335-77-3	P070825026	ND	0.278	0.510			U	
4:2 FTS		757124-72-4	P070825026	ND	0.0686	0.496			U	
6:2 FTS		27619-97-2	P070825026	ND	0.250	0.504			U	
8:2 FTS		39108-34-4	P070825026	ND	0.119	0.507			U	
10:2 FTS	120226-60-0	P070825026	ND	0.405	0.529			U		
Sulfonamidos	FBSA	30334-69-1	P070825026	0.347	0.251	0.529			J	
	N-EiFOSA	4151-50-2	P070825026	ND	0.327	0.529			U	
	N-EiFOSAA	2991-50-6	P070825026	0.0829	0.215	0.529			L	
	N-EiFOSE	1691-99-2	P070825026	ND	0.810	2.38			U	
	N-MeFOSA	31506-32-8	P070825026	ND	0.218	0.529			U	
	N-MeFOSAA	2355-31-9	P070825026	0.00101	0.149	0.529			L	
	N-MeFOSE	24448-09-7	P070825026	ND	0.503	2.38			U	
	PFOSA	754-91-6	P070825026	0.0456	0.0743	0.529			L	
	PFECAs	ADONA	919005-14-4	P070825026	ND	0.143	0.501			U
EVE Acid		69087-46-3	P070825026	ND	0.169	1.19			U	
HFPO-DA		13252-13-6	P070825026	1.61	0.0561	0.529			U	
Hydro-EVE Acid		773804-62-9	P070825026	ND	0.174	0.529			U	
NFDHA		151772-58-6	P070825026	ND	0.111	0.529			U	
PEPA		267239-61-2	P070825026	1.10	0.0992	0.529			U	
PFECA-G		801212-59-9	P070825026	ND	0.0706	0.529			U	
PFMOAA		674-13-5	P070825026	13.0	0.268	0.529			U	
PFMOBA		863090-89-5	P070825026	ND	0.888	1.19			U	
PFMOPrA		377-73-1	P070825026	ND	0.189	0.529			U	
PFO2HxA		39492-88-1	P070825026	3.54	0.170	0.529			U	
PFO3OA		39492-89-2	P070825026	1.31	0.243	0.529			U	
PFO4DA		39492-90-5	P070825026	ND	0.418	2.65			U	
PFO5DA		39492-91-6	P070825026	ND	0.423	2.65			U	
PMPA		13140-29-9	P070825026	2.22	0.125	0.529			U	
R-EVE		2416366-22-6	P070825026	2.28	0.878	1.19			U	
PFESAs		11Cl-PF3OUdS	763051-92-9	P070825026	ND	0.250	0.499			U
		9Cl-PF3ONS	756426-58-1	P070825026	ND	0.339	0.493			U
	Hydrolyzed PSDA	2416366-19-1	P070825026	1.26	0.352	0.529			U	
	Nafion Byproduct 1 (PS Acid)	29311-67-9	P070825026	ND	0.283	0.529			U	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	P070825026	0.204	0.438	0.529			L	
	NVHOS	1132933-86-8	P070825026	1.36	0.0815	0.529			U	
	PFEESA	113507-82-7	P070825026	ND	0.159	0.529			U	
	R-PSDA	2416366-18-0	P070825026	3.74	2.33	2.33			U	
	R-PSDCA	2416366-21-5	P070825026	ND	0.223	0.529			U	
ES	MPFBA		P070825026				20-150%	71.7%		
	M5PFPeA		P070825026				20-150%	185%	Q	
	M3PFBS		P070825026				20-150%	230%	Q	
	M2-4:2 FTS		P070825026				20-150%	108%		
	M5PFFhxA		P070825026				20-150%	92.1%		
	M3HFPO-DA		P070825026				20-150%	87.1%		
	M4PFFHpA		P070825026				20-150%	89.4%		

Enthalpy Analytical

Job No.: 0825-736-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	080125-SO1		
Sampling Site			
Enthalpy ID	0825-736-001-2	Prep Batch	EU118528
Matrix	aqueous	Analyst	jogres
Sampling Date	2025-08-01 13:35	Instrument	Pippin
Received Date	2025-08-01	Sample Vol mL	302.32
Prep Date	2025-08-06 09:05	Extract Vol mL	0.4
AnalysisDate	2025-08-08 01:12	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	A		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M3PFHxS		P070825026				20-150%	81.2%	
M2-6:2 FTS		P070825026				20-150%	76.0%	
M8PFOA		P070825026				20-150%	87.3%	
M9PFNA		P070825026				20-150%	93.2%	
M8PFOS		P070825026				20-150%	89.1%	
M2-8:2 FTS		P070825026				20-150%	69.5%	
M8FOSA-I		P070825026				20-150%	83.1%	
M6PFDA		P070825026				20-150%	85.1%	
d3-N-MeFOSAA		P070825026				20-150%	79.7%	
d5-N-EtFOSAA		P070825026				20-150%	83.5%	
M7PFUdA		P070825026				20-150%	86.2%	
MPFDoA		P070825026				20-150%	75.2%	
M2PFTeDA		P070825026				20-150%	56.2%	
d3-N-MeFOSA		P070825026				10-200%	38.2%	
d5-N-EtFOSA		P070825026				10-200%	30.9%	
d7-N-MeFOSE		P070825026				10-200%	63.0%	
d9-N-EtFOSE		P070825026				10-200%	56.7%	

Enthalpy Analytical

Job No.: 0825-736-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	080125-EO1	Prep Batch	EU118522
Sampling Site		Analyst	jogres
Enthalpy ID	0825-736-002-1	Instrument	Bumblebee
Matrix	aqueous	Sample Vol mL	0.1
Sampling Date	2025-08-01 13:35	Extract Vol mL	0.2
Received Date	2025-08-01	Split Factor	N/A
Prep Date	2025-08-05 12:30	Method Code	EU-047-NPW
AnalysisDate	2025-08-05 22:36		
SampleType	Sample		
Bottle ID	A		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	8050825-08052236	419	700	1530			L
ES	13C3-PFPrA		8050825-08052236				20-150%	91.9%	

Enthalpy Analytical

Job No.: 0825-736-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name 080125-EO1
 Sampling Site
 Enthalpy ID 0825-736-002-2 Prep Batch EU118528
 Matrix aqueous Analyst jogres
 Sampling Date 2025-08-01 13:35 Instrument Pippin
 Received Date 2025-08-01 Sample Vol mL 308.4
 Prep Date 2025-08-06 09:05 Extract Vol mL 0.4
 AnalysisDate 2025-08-08 01:34 Split Factor N/A
 SampleType Sample Method Code EU-047-NPW
 Bottle ID A

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	P070825027	4.29	0.206	0.519				
	PFPeA	2706-90-3	P070825027	5.80	0.148	0.519				
	PFHxA	307-24-4	P070825027	5.06	0.173	0.519				
	PFFHpA	375-85-9	P070825027	2.75	0.182	0.519				
	PFOA	335-67-1	P070825027	6.36	0.119	0.519				
	PFNA	375-95-1	P070825027	1.00	0.117	0.519				
	PFDA	335-76-2	P070825027	0.702	0.148	0.519				
	PFUnDA	2058-94-8	P070825027	0.118	0.117	0.519			J	
	PFDoDA	307-55-1	P070825027	ND	0.211	0.519			U	
	PFTrDA	72629-94-8	P070825027	ND	0.172	0.519			U	
	PFTeDA	376-06-7	P070825027	ND	0.198	0.519			U	
	PFHxDA	67905-19-5	P070825027	ND	0.276	0.519			U	
	Sulfonates	PFBS	375-73-5	P070825027	2.71	0.276	0.519			
		PFPeS	2706-91-4	P070825027	0.501	0.107	0.489			
PFHxS		355-46-4	P070825027	3.17	0.400	0.475				
PFFHpS		375-92-8	P070825027	0.0967	0.251	0.494			L	
PFOS		1763-23-1	P070825027	13.2	0.274	0.481				
PFNS		68259-12-1	P070825027	ND	0.161	0.500			U	
PFDS		335-77-3	P070825027	ND	0.272	0.500			U	
4:2 FTS		757124-72-4	P070825027	ND	0.0673	0.486			U	
6:2 FTS		27619-97-2	P070825027	ND	0.245	0.494			U	
8:2 FTS		39108-34-4	P070825027	0.00322	0.116	0.497			L	
10:2 FTS	120226-60-0	P070825027	ND	0.397	0.519			U		
Sulfonamidos	FBSA	30334-69-1	P070825027	0.299	0.246	0.519			J	
	N-EiFOSA	4151-50-2	P070825027	ND	0.321	0.519			U	
	N-EiFOSAA	2991-50-6	P070825027	0.0441	0.211	0.519			L	
	N-EiFOSE	1691-99-2	P070825027	ND	0.794	2.33			U	
	N-MeFOSA	31506-32-8	P070825027	ND	0.214	0.519			U	
	N-MeFOSAA	2355-31-9	P070825027	ND	0.146	0.519			U	
	N-MeFOSE	24448-09-7	P070825027	ND	0.493	2.33			U	
	PFOSA	754-91-6	P070825027	ND	0.0728	0.519			U	
	PFECAs	ADONA	919005-14-4	P070825027	ND	0.141	0.492			U
EVE Acid		69087-46-3	P070825027	ND	0.165	1.17			U	
HFPO-DA		13252-13-6	P070825027	1.58	0.0550	0.519			U	
Hydro-EVE Acid		773804-62-9	P070825027	ND	0.170	0.519			U	
NFDHA		151772-58-6	P070825027	ND	0.109	0.519			U	
PEPA		267239-61-2	P070825027	1.04	0.0973	0.519			U	
PFECA-G		801212-59-9	P070825027	ND	0.0692	0.519			U	
PFMOAA		674-13-5	P070825027	10.9	0.263	0.519			U	
PFMOBA		863090-89-5	P070825027	ND	0.871	1.17			U	
PFMOPrA		377-73-1	P070825027	ND	0.185	0.519			U	
PFO2HxA		39492-88-1	P070825027	3.31	0.167	0.519			U	
PFO3OA		39492-89-2	P070825027	1.13	0.238	0.519			U	
PFO4DA		39492-90-5	P070825027	ND	0.410	2.59			U	
PFO5DA		39492-91-6	P070825027	ND	0.415	2.59			U	
PMPA		13140-29-9	P070825027	2.27	0.122	0.519			U	
R-EVE		2416366-22-6	P070825027	1.99	0.861	1.17			U	
PFESAs		11Cl-PF3OUdS	763051-92-9	P070825027	ND	0.245	0.489			U
		9Cl-PF3ONS	756426-58-1	P070825027	ND	0.332	0.483			U
		Hydrolyzed PSDA	2416366-19-1	P070825027	1.44	0.345	0.519			U
	Nafion Byproduct 1 (PS Acid)	29311-67-9	P070825027	ND	0.277	0.519			U	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	P070825027	0.233	0.430	0.519			L	
	NVHOS	1132933-86-8	P070825027	1.64	0.0799	0.519			U	
	PFEESA	113507-82-7	P070825027	ND	0.156	0.519			U	
	R-PSDA	2416366-18-0	P070825027	4.34	2.29	2.29			U	
ES	R-PSDCA	2416366-21-5	P070825027	ND	0.219	0.519			U	
	MPFBA		P070825027				20-150%	89.0%		
	M5PFPeA		P070825027				20-150%	180%	Q	
	M3PFBS		P070825027				20-150%	264%	Q	
	M2-4:2 FTS		P070825027				20-150%	143%		
	M5PFFHpA		P070825027				20-150%	98.0%		
	M3HFPO-DA		P070825027				20-150%	82.2%		
M4PFFHpA		P070825027				20-150%	97.2%			

Enthalpy Analytical

Job No.: 0825-736-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	080125-EO1	Prep Batch	EU118528
Sampling Site		Analyst	jogres
Enthalpy ID	0825-736-002-2	Instrument	Pippin
Matrix	aqueous	Sample Vol mL	308.4
Sampling Date	2025-08-01 13:35	Extract Vol mL	0.4
Received Date	2025-08-01	Split Factor	N/A
Prep Date	2025-08-06 09:05	Method Code	EU-047-NPW
AnalysisDate	2025-08-08 01:34		
SampleType	Sample		
Bottle ID	A		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M3PFHxS		P070825027				20-150%	94.8%	
M2-6:2 FTS		P070825027				20-150%	100%	
M8PFOA		P070825027				20-150%	91.2%	
M9PFNA		P070825027				20-150%	99.2%	
M8PFOS		P070825027				20-150%	98.6%	
M2-8:2 FTS		P070825027				20-150%	91.1%	
M8FOSA-I		P070825027				20-150%	98.8%	
M6PFDA		P070825027				20-150%	91.5%	
d3-N-MeFOSAA		P070825027				20-150%	97.4%	
d5-N-EtFOSAA		P070825027				20-150%	100%	
M7PFUdA		P070825027				20-150%	97.5%	
MPFDoA		P070825027				20-150%	86.2%	
M2PFTeDA		P070825027				20-150%	78.0%	
d3-N-MeFOSA		P070825027				10-200%	58.0%	
d5-N-EtFOSA		P070825027				10-200%	48.4%	
d7-N-MeFOSE		P070825027				10-200%	87.6%	
d9-N-EtFOSE		P070825027				10-200%	89.6%	

QC Data

Enthalpy Analytical

Job No.: 0825-736-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	MB_118522_PFAS	Prep Batch	EU118522
Sampling Site		Analyst	jogres
Enthalpy ID	MB_118522_PFAS	Instrument	Bumblebee
Matrix	aqueous	Sample Vol mL	0.1
Sampling Date		Extract Vol mL	0.2
Received Date		Split Factor	N/A
Prep Date	2025-08-05 12:30	Method Code	EU-047-NPW
AnalysisDate	2025-08-05 20:52		
SampleType	Blank		
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	8050825-08052052	409	700	1530			L
ES	13C3-PFPrA		8050825-08052052				20-150%	90.0%	

Enthalpy Analytical

Job No.: 0825-736-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	MB_118528_PFAS	Prep Batch	EU118528
Sampling Site		Analyst	jogres
Enthalpy ID	MB_118528_PFAS	Instrument	Pippin
Matrix	aqueous	Sample Vol mL	250
Sampling Date		Extract Vol mL	0.4
Received Date		Split Factor	N/A
Prep Date	2025-08-06 09:05	Method Code	EU-047-NPW
AnalysisDate	2025-08-07 18:23		
SampleType	Blank		
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	P070825008	ND	0.254	0.640			U	
	PFPeA	2706-90-3	P070825008	ND	0.183	0.640			U	
	PFFhxA	307-24-4	P070825008	ND	0.214	0.640			U	
	PFFHpA	375-85-9	P070825008	ND	0.224	0.640			U	
	PFOA	335-67-1	P070825008	ND	0.146	0.640			U	
	PFNA	375-95-1	P070825008	ND	0.145	0.640			U	
	PFDA	335-76-2	P070825008	ND	0.183	0.640			U	
	PFUnDA	2058-94-8	P070825008	ND	0.145	0.640			U	
	PFFDoDA	307-55-1	P070825008	ND	0.260	0.640			U	
	PFTTrDA	72629-94-8	P070825008	ND	0.212	0.640			U	
	PFTeDA	376-06-7	P070825008	ND	0.244	0.640			U	
	PFFhxDA	67905-19-5	P070825008	ND	0.340	0.640			U	
	Sulfonates	PFBS	375-73-5	P070825008	ND	0.340	0.640			U
		PFPeS	2706-91-4	P070825008	ND	0.131	0.603			U
		PFFhXS	355-46-4	P070825008	ND	0.494	0.586			U
PFFHpS		375-92-8	P070825008	ND	0.310	0.610			U	
PFOs		1763-23-1	P070825008	ND	0.338	0.593			U	
PFNS		68259-12-1	P070825008	ND	0.199	0.616			U	
PFDS		335-77-3	P070825008	ND	0.336	0.616			U	
4:2 FTS		757124-72-4	P070825008	ND	0.0830	0.600			U	
6:2 FTS		27619-97-2	P070825008	ND	0.302	0.610			U	
8:2 FTS		39108-34-4	P070825008	ND	0.143	0.613			U	
10:2 FTS	120226-60-0	P070825008	ND	0.490	0.640			U		
Sulfonamidos	FBSA	30334-69-1	P070825008	ND	0.304	0.640			U	
	N-EiFOSA	4151-50-2	P070825008	ND	0.396	0.640			U	
	N-EiFOSAA	2991-50-6	P070825008	ND	0.260	0.640			U	
	N-EiFOSE	1691-99-2	P070825008	ND	0.980	2.88			U	
	N-MeFOSA	31506-32-8	P070825008	ND	0.264	0.640			U	
	N-MeFOSAA	2355-31-9	P070825008	ND	0.180	0.640			U	
	N-MeFOSE	24448-09-7	P070825008	ND	0.608	2.88			U	
	PFOsA	754-91-6	P070825008	ND	0.0898	0.640			U	
	PFECAs	ADONA	919005-14-4	P070825008	ND	0.173	0.606			U
EVE Acid		69087-46-3	P070825008	ND	0.204	1.44			U	
HFPO-DA		13252-13-6	P070825008	ND	0.0678	0.640			U	
Hydro-EVE Acid		773804-62-9	P070825008	ND	0.210	0.640			U	
NFDHA		151772-58-6	P070825008	0.00906	0.135	0.640			L	
PEPA		267239-61-2	P070825008	ND	0.120	0.640			U	
PFECA-G		801212-59-9	P070825008	ND	0.0854	0.640			U	
PFMOAA		674-13-5	P070825008	ND	0.324	0.640			U	
PFMOBA		863090-89-5	P070825008	ND	1.07	1.44			U	
PFMOPrA		377-73-1	P070825008	ND	0.228	0.640			U	
PFO2HxA		39492-88-1	P070825008	ND	0.206	0.640			U	
PFO3OA		39492-89-2	P070825008	ND	0.294	0.640			U	
PFO4DA		39492-90-5	P070825008	ND	0.506	3.20			U	
PFO5DA		39492-91-6	P070825008	ND	0.512	3.20			U	
PMPA		13140-29-9	P070825008	ND	0.151	0.640			U	
R-EVE		2416366-22-6	P070825008	ND	1.06	1.44			U	
PFESAs		11CI-PF3OUdS	763051-92-9	P070825008	ND	0.302	0.603			U
		9CI-PF3ONS	756426-58-1	P070825008	ND	0.410	0.596			U
	Hydrolyzed PSDA	2416366-19-1	P070825008	ND	0.426	0.640			U	
	Nafion Byproduct 1 (PS Acid)	29311-67-9	P070825008	ND	0.342	0.640			U	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	P070825008	ND	0.530	0.640			U	
	NVHOS	1132933-86-8	P070825008	ND	0.0986	0.640			U	
	PFEESA	113507-82-7	P070825008	ND	0.192	0.640			U	
	R-PSDA	2416366-18-0	P070825008	ND	2.82	2.82			U	
	R-PSDCA	2416366-21-5	P070825008	ND	0.270	0.640			U	
	ES	MPFBA		P070825008				20-150%	90.1%	
M5PFPeA			P070825008				20-150%	98.9%		
M3PFBS			P070825008				20-150%	94.4%		
M2-4:2 FTS			P070825008				20-150%	95.4%		
M5PFFhxA			P070825008				20-150%	97.4%		
M3HFPO-DA			P070825008				20-150%	94.7%		
M4PFFHpA			P070825008				20-150%	88.5%		

Enthalpy Analytical

Job No.: 0825-736-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	MB_118528_PFAS	Prep Batch	EU118528
Sampling Site		Analyst	jogres
Enthalpy ID	MB_118528_PFAS	Instrument	Pippin
Matrix	aqueous	Sample Vol mL	250
Sampling Date		Extract Vol mL	0.4
Received Date		Split Factor	N/A
Prep Date	2025-08-06 09:05	Method Code	EU-047-NPW
AnalysisDate	2025-08-07 18:23		
SampleType	Blank		
Bottle ID	-		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M3PFHxS		P070825008				20-150%	83.3%	
M2-6:2 FTS		P070825008				20-150%	68.1%	
M8PFOA		P070825008				20-150%	85.7%	
M9PFNA		P070825008				20-150%	78.2%	
M8PFOS		P070825008				20-150%	78.1%	
M2-8:2 FTS		P070825008				20-150%	54.8%	
M8FOSA-I		P070825008				20-150%	72.8%	
M6PFDA		P070825008				20-150%	66.4%	
d3-N-MeFOSAA		P070825008				20-150%	64.7%	
d5-N-EtFOSAA		P070825008				20-150%	64.2%	
M7PFUdA		P070825008				20-150%	60.7%	
MPFDoA		P070825008				20-150%	54.5%	
M2PFTeDA		P070825008				20-150%	56.0%	
d3-N-MeFOSA		P070825008				10-200%	23.3%	
d5-N-EtFOSA		P070825008				10-200%	18.8%	
d7-N-MeFOSE		P070825008				10-200%	53.6%	
d9-N-EtFOSE		P070825008				10-200%	54.3%	

Enthalpy Analytical

Job No.: 0825-736-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Enthalpy ID	OPR_118522_PFAS	Prep Batch	EU118522	Sample Vol (mL)	0.1
Sample Name	OPR_118522_PFAS	Prep Date	2025-08-05 12:30	Extract Vol (mL)	0.2
Matrix	aqueous	Analysis Date	2025-08-05 21:03	Split Factor	N/A
Sampling Date		Analyst	jogres	Method Code	EU-047-NPW
Received Date		Instrument	Bumblebee	Sample Type	Control
		Bottle ID	-		

	Compound	CAS	InjFileName	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	B050825-08052103	26600	700	1530	40-150%	133%	
ES	¹³ C3-PFPrA		B050825-08052103				20-150%	73.8%	

Enthalpy Analytical

Job No.: 0825-736-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Enthalpy ID	OPR_118528_PFAS	Prep Batch	EU118528	Sample Vol (mL)	250
Sample Name	OPR_118528_PFAS	Prep Date	2025-08-06 09:05	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2025-08-07 18:45	Split Factor	N/A
Sampling Date		Analyst	jogres	Method Code	EU-047-NPW
Received Date		Instrument	Pippin	Sample Type	Control
		Bottle ID	-		

	Compound	CAS	InjFileName	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	P070825009	19.9	0.254	0.640	47.9-144%	99.7%		
	PFPeA	2706-90-3	P070825009	19.4	0.183	0.640	41.7-159%	96.9%		
	PFHxA	307-24-4	P070825009	20.8	0.214	0.640	43.2-154%	104%		
	PFHpA	375-85-9	P070825009	21.9	0.224	0.640	42.1-155%	109%		
	PFOA	335-67-1	P070825009	19.3	0.146	0.640	51.1-148%	96.7%		
	PFNA	375-95-1	P070825009	21.1	0.145	0.640	51.6-153%	105%		
	PFDA	335-76-2	P070825009	21.2	0.183	0.640	44.5-156%	106%		
	PFUnDA	2058-94-8	P070825009	20.0	0.145	0.640	40.3-156%	99.8%		
	PFDoDA	307-55-1	P070825009	20.1	0.260	0.640	40.4-158%	101%		
	PFTeDA	72629-04-8	P070825009	23.0	0.212	0.640	42.2-201%	115%		
	PFTeDA	376-06-7	P070825009	21.5	0.244	0.640	43-162%	107%		
	Sulfonates	PFBS	375-73-5	P070825009	18.3	0.340	0.640	42.7-155%	103%	
		PFPeS	2706-91-4	P070825009	17.8	0.131	0.603	40.3-152%	94.5%	
		PFHxS	355-46-4	P070825009	18.3	0.494	0.586	45-148%	100%	
PFHpS		375-92-8	P070825009	19.6	0.310	0.610	39.8-166%	103%		
PFOS		1763-23-1	P070825009	18.3	0.338	0.593	59.2-132%	98.6%		
PFNS		68259-12-1	P070825009	19.3	0.199	0.616	38.1-153%	100%		
PFDS		335-77-3	P070825009	18.1	0.336	0.616	28.6-148%	93.9%		
4:2 FTS		757124-72-4	P070825009	19.6	0.0830	0.600	41.5-157%	105%		
6:2 FTS		27619-97-2	P070825009	20.6	0.302	0.610	44.5-160%	108%		
8:2 FTS		39108-34-4	P070825009	20.0	0.143	0.613	39.4-166%	104%		
Sulfonamidos		N-EtFOSA	4151-50-2	P070825009	25.4	0.396	0.640	26.7-172%	127%	
		N-EtFOSAA	2991-50-6	P070825009	20.1	0.260	0.640	42.8-156%	101%	
	N-EtFOSE	1691-99-2	P070825009	113	0.980	2.88	38.9-161%	125%		
	N-MeFOSA	31506-32-8	P070825009	27.2	0.264	0.640	26.4-183%	136%		
	N-MeFOSAA	2355-31-9	P070825009	21.0	0.180	0.640	42-155%	105%		
	N-MeFOSE	24448-09-7	P070825009	119	0.608	2.88	37.6-155%	132%		
	PFOSA	754-91-6	P070825009	20.0	0.0898	0.640	39.1-158%	100%		
PFECAs	ADONA	919005-14-4	P070825009	20.7	0.173	0.606	32.2-151%	103%		
	HFPO-DA	13252-13-6	P070825009	21.0	0.0678	0.640	61.8-131%	105%		
PFESAs	11Cl-PF3OUdS	763051-92-9	P070825009	17.5	0.302	0.603	21.8-141%	87.5%		
	9Cl-PF3ONS	756426-58-1	P070825009	18.4	0.410	0.596	37.6-146%	92.2%		
ES	MPFBA		P070825009				20-150%	94.7%		
	M5PFPeA		P070825009				20-150%	102%		
	M3PFBS		P070825009				20-150%	95.8%		
	M2-4:2 FTS		P070825009				20-150%	89.0%		
	M5PFHxA		P070825009				20-150%	101%		
	M3HFPO-DA		P070825009				20-150%	104%		
	M4PFHpA		P070825009				20-150%	92.0%		
	M3PFHxS		P070825009				20-150%	93.6%		
	M2-6:2 FTS		P070825009				20-150%	69.2%		
	M8PFOA		P070825009				20-150%	95.2%		
	M9PFNA		P070825009				20-150%	92.5%		
	M8PFOS		P070825009				20-150%	94.1%		
	M2-8:2 FTS		P070825009				20-150%	69.3%		
	M8FOSA-I		P070825009				20-150%	83.2%		
	M6PFDA		P070825009				20-150%	89.8%		
	d3-N-MeFOSAA		P070825009				20-150%	84.3%		
	d5-N-EtFOSAA		P070825009				20-150%	83.5%		
	M7PFUdA		P070825009				20-150%	90.2%		
	MPFDoA		P070825009				20-150%	88.6%		
	M2PFTeDA		P070825009				20-150%	77.0%		
	d3-N-MeFOSA		P070825009				10-200%	15.9%		
	d5-N-EtFOSA		P070825009				10-200%	15.8%		
	d7-N-MeFOSE		P070825009				10-200%	73.1%		
d9-N-EtFOSE		P070825009				10-200%	72.5%			

Sample Custody



JOB ID: Date / Time: 8/1/25 14:46 Initials: SBS
 OR
 Client: A Brunswick

TE SBS 8/1/25

Cooler of

Temp °C: 5.6 Thermometer ID: 716

Received via FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input checked="" type="checkbox"/> Other <input type="checkbox"/>	<i>Check one</i> On ice: <input checked="" type="checkbox"/> Melted ice: <input type="checkbox"/> Ambient: <input type="checkbox"/>	<i>Check one</i> in a Box: <input type="checkbox"/> in a Cooler: <input checked="" type="checkbox"/> Cooler in Box: <input type="checkbox"/>	<table border="0"> <tr> <td></td> <td>Yes</td> <td>No</td> </tr> <tr> <td>Cooler seals:</td> <td><input type="checkbox"/></td> <td><input checked="" type="checkbox"/></td> </tr> <tr> <td>Sample seals:</td> <td><input type="checkbox"/></td> <td><input checked="" type="checkbox"/></td> </tr> <tr> <td>Good condition:</td> <td><input checked="" type="checkbox"/></td> <td><input type="checkbox"/></td> </tr> </table>		Yes	No	Cooler seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Sample seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Good condition:	<input checked="" type="checkbox"/>	<input type="checkbox"/>
		Yes	No												
	Cooler seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>												
	Sample seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>												
	Good condition:	<input checked="" type="checkbox"/>	<input type="checkbox"/>												
	Comment: <input type="text"/>														

Temp °C: Thermometer ID: Cooler of

Received via FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input type="checkbox"/> Other <input type="checkbox"/>	<i>Check one</i> On ice: <input type="checkbox"/> Melted ice: <input type="checkbox"/> Ambient: <input type="checkbox"/>	<i>Check one</i> in a Box: <input type="checkbox"/> in a Cooler: <input type="checkbox"/> Cooler in Box: <input type="checkbox"/>	<table border="0"> <tr> <td></td> <td>Yes</td> <td>No</td> </tr> <tr> <td>Cooler seals:</td> <td><input type="checkbox"/></td> <td><input type="checkbox"/></td> </tr> <tr> <td>Sample seals:</td> <td><input type="checkbox"/></td> <td><input type="checkbox"/></td> </tr> <tr> <td>Good condition:</td> <td><input type="checkbox"/></td> <td><input type="checkbox"/></td> </tr> </table>		Yes	No	Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>	Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>	Good condition:	<input type="checkbox"/>	<input type="checkbox"/>
		Yes	No												
	Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>												
	Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>												
	Good condition:	<input type="checkbox"/>	<input type="checkbox"/>												
	Comment: <input type="text"/>														

Temp °C: Thermometer ID: Cooler of

Received via FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input type="checkbox"/> Other <input type="checkbox"/>	<i>Check one</i> On ice: <input type="checkbox"/> Melted ice: <input type="checkbox"/> Ambient: <input type="checkbox"/>	<i>Check one</i> in a Box: <input type="checkbox"/> in a Cooler: <input type="checkbox"/> Cooler in Box: <input type="checkbox"/>	<table border="0"> <tr> <td></td> <td>Yes</td> <td>No</td> </tr> <tr> <td>Cooler seals:</td> <td><input type="checkbox"/></td> <td><input type="checkbox"/></td> </tr> <tr> <td>Sample seals:</td> <td><input type="checkbox"/></td> <td><input type="checkbox"/></td> </tr> <tr> <td>Good condition:</td> <td><input type="checkbox"/></td> <td><input type="checkbox"/></td> </tr> </table>		Yes	No	Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>	Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>	Good condition:	<input type="checkbox"/>	<input type="checkbox"/>
		Yes	No												
	Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>												
	Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>												
	Good condition:	<input type="checkbox"/>	<input type="checkbox"/>												
	Comment: <input type="text"/>														