

County of Brunswick

3954 Clearwell Dr NE
Leland, NC 28451

Northwest Plant

Leland, NC
Samples Received: 3-7-19

Analytical Report (0319-707)

Isotope Dilution PFAS



Enthalpy Analytical, LLC – Ultratrace

Phone: (910) 212-5858 / www.enthalpy.com
2714 Exchange Drive, Wilmington, NC 28405

I certify that to the best of my knowledge all analytical data presented in this report:

- Have been checked for completeness
- Are accurate, error-free, and legible
- Have been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s)

This analytical report was prepared in Portable Document Format (.PDF) and contains _____ pages.

....."Report Issued Date: _____



Summary of Results

Summary of Results: PFAS
Enthalpy Ultratrace Batch #
10237
PFAS

Analyte	Method Blank ng/L	030719-S01 ng/L	030719-E01 ng/L
Acids			
PFBA	2.67	3.68 B	2.72 B
PFPeA	<0.898 U	<0.633 U	3.27
PFHxA	<1.58 U	2.85	3.50
PFHpA	<0.695 U	1.83	1.78
PFOA	<0.795 U	3.05	2.94
PFNA	<0.509 U	<0.359 U	<0.359 U
PFDA	<1.25 U	<0.882 U	<0.882 U
PFUnA	<0.481 U	<0.339 U	<0.339 U
PFDoA	<0.475 U	<0.335 U	<0.335 U
PFTTrA	<0.745 U	<0.525 U	<0.526 U
PFTA	<0.830 U	<0.585 U	<0.586 U
Sulfonates			
L-PFBS	<0.830 U	<0.585 U	<0.586 U
PFPeS	<0.990 U	<0.698 U	<0.699 U
PFHxS	<0.827 U	2.81	3.99
PFHpS	<0.779 U	<0.549 U	<0.550 U
PFOS	<0.471 U	8.07	7.87
PFNS	<0.654 U	<0.461 U	<0.462 U
PFDS	<1.35 U	<0.952 U	<0.953 U
4:2 FTS	<0.646 U	<0.456 U	<0.456 U
6:2 FTS	<0.723 U	<0.510 U	<0.510 U
8:2 FTS	<0.569 U	<0.401 U	<0.402 U
Other			
PFOSA	<3.65 U	<2.57 U	<2.58 U
N-MeFOSAA	<0.544 U	<0.384 U	<0.384 U
N-EtFOSAA	<0.651 U	<0.459 U	<0.459 U
HFPO-DA (Gen-X)	<33.3 U	<23.5 U	<23.5 U
Lab Sample ID	MB_10237	0319-707_10237_001	0319-707_10237_002

""""HpcnParrative

Enthalpy Analytical Narrative Summary

Company	County of Brunswick
Job#	0319-707 PFAS
Client Project #	n/a

Custody	<p>Braidy May of Enthalpy Analytical Wilmington received the samples (via client courier) on 03/07/19 on ice at 4.8°C in good condition.</p> <p>Prior to, during, and after analysis, the samples were stored in the laboratory with access only by authorized personnel of Enthalpy Analytical, LLC.</p>
Analysis	<p>The samples were analyzed by isotope dilution method for PFAS using Waters Acquity UPLC equipped with Xevo TQ MS (LC/MS/MS “Fili”).</p> <p>For each sample, a 35mL aliquot was weighed and spiked with Extraction Standard (ES). The sample was then mixed well.</p> <p>Cleanup procedures were performed on the supernatant and then extracted via SPE. Each final sample extract was transferred to an autosampler vial and spiked with 200 µL of Injection Standard (IS), prior to analysis.</p>
Calibration	<p>The analytes and labeled standards in the initial calibration exhibited RSDs less than 50%. All analytes passed the R² coefficient correlation criteria. The continuing calibration met the ±30% criteria for native analytes, ±50% criteria for ES recoveries, and ±50% of IS peak area.</p>
QC Notes	<p>The LCS injection met the ±50% criteria for ES recoveries and IS Area with the following exceptions: M8FOSA and MPFOS fell outside the lower limits for ES recovery and IS area count, respectively.</p> <p>PFBA was detected in the method blank greater than 1/10 the sample amount. This analyte detected in the samples is notated with a B qualifier. Analytes, M3PFHxS and M8FOSA fell outside the lower limits for ES recovery. Additionally, MPFDA and MPFOS fell outside the lower limits for IS area count.</p> <p>The samples were extracted within the 14-day from collection holding time. Extracts were analyzed within the 28-days from extraction to analysis holding time required by the method.</p>
Reporting Notes	<p>ES recoveries in sample 030719-S01 fell outside the lower limit for M8FOSA and M2PFTeDA. ES recoveries greater than the upper limit for M3HFPO-DA were also observed.</p>



Enthalpy Analytical Narrative Summary (continued)

Reporting Notes (continued)

ES recoveries in sample 030719-E01 fell outside the lower limit for M8FOSA. ES recoveries greater than the upper limit for M3HFPO-DA were also observed.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

The results presented in this report are representative of the samples as provided to the laboratory.

The samples, calibrations and standards for the data presented in this report were analyzed at 2714 Exchange Drive, Wilmington, NC 28405.



General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, Inc.-Wilmington, NC data reports, unless specifically noted otherwise.

General Data Qualifiers / Data Attributes

- B – The analyte was found in the method blank, at a concentration that was at least 10% of the concentration in the sample.
- C – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve).
- EMPC – Represents an estimated maximum possible concentration. EMPCs arise in cases where the signal/noise ratio is not sufficient for peak identification (the determined ion-abundance ratio is outside the allowed theoretical range), or where there is a co-eluting interference.
- J – Indicates that an analyte has a concentration below the reporting limit (lowest point of the calibration curve).
- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable.
- PR – Due to interference, the associated congener is poorly resolved.
- DI – Indicates the presence of a quantitative interference.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected. The Estimated Detection Limit (EDL) may be reported for this analyte.
- V – The labeled standard recovery was found to be outside of the method control limits.

DRBC/TMDL Specific Data Qualifiers / Data Attributes

- J – The reported result is an estimate. The value is less than the minimum calibration level but greater than the Estimated Detection Limit (EDL).
- U – The analyte was not detected in the sample at the Estimated Detection Limit (EDL).
- E – The reported concentration is an estimate. The value exceeds the upper calibration range (upper point of the calibration curve).




General Reporting Notes – Data Qualifiers

- D – Dilution Data. Result was obtained from the analysis of a dilution.
- B – Analyte found in the sample and associated method blank.
- Cxx – Co-elutes with the indicated congener, data is reported under the lowest IUPAC congener. ‘xx’ denotes the IUPAC number with the lowest numerical designated congener.
- NR – Analyte is not reportable because of problems in sample preparation or analysis.
- V – Labeled standard recovery is not within method control limits.
- X – Results from re-injection/repeat/second-column analysis.
- EMPC – Estimated Maximum Possible Concentration. Indicates that a peak is identified but did not meet the method specified ion-abundance ratio.

Lab Identifiers

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Indicates a dilution of the sample extract. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.
- RJ – Indicates a reinjection of the sample extract.
- S – Indicates a sample split. The number that follows the “S” indicates the split factor.

Job 0319-707 report due Thursday Mar 21 2019



To	Glenn Walker	From	Lindsay Boone
Company	County of Brunswick	Customer PO	
Site	N.W. Leland Plant	Sales Order	SO1035982
Customer Phone	910-612-5618	Date Due QA	Tuesday, March 19, 2019
Customer Email	glenn.walker@brunswickcountync.gov	Results Due	Thursday, March 21, 2019
Customer Project		Report Due	Thursday, March 21, 2019

Line	Method	Analytes	Quant	Dept	Objective	Status
1	WM-026 PFAS by Isotope Dilution (non-potable water) report:	Legacy 24 plus Gen X	2	Wilmington	Level 1	Sampled 03-07-19 10:55 (local time) Received 03-07-19 12:45 ReceivedTemp 4.8 StorageTemp 4.0 StorageLocation R1

**This Is The Last Page
Of This Report.**